

Photometry Analysis Manual

pHotoFlex / pHotoFlex Turb



Practical part and analysis specifications

Prog. V 1.20 W

**Accuracy when
going to press**

The use of advanced technology and the high quality standard of our instruments are the result of continuous development. This may result in differences between this operating manual and your instrument. Also, we cannot guarantee that there are absolutely no errors in this manual. Therefore, we are sure you will understand that we cannot accept any legal claims resulting from the data, figures or descriptions.

**Up-to-dateness
of firmware**

Part of the process of consequently improving our products is the continuous further development of instrument firmware. The current firmware for the pHotoFlex (Turb) can be found on the Internet. You can easily transfer it to your instrument with the aid of the AK 540/B cable provided and a Personal Computer.

More detailed information can be found in the appendix of this operating manual or on the Internet under <http://www.WTW.com>.

Copyright

© 2010, WTW GmbH

Reprinting - even as excerpts - is only allowed with the explicit written authorization of WTW GmbH, Weilheim.

Printed in Germany.

| | | |
|----------|--|-----------|
| 1 | Practical part | 7 |
| 1.1 | For your safety | 7 |
| 1.2 | General information on test sets | 7 |
| 1.3 | Carrying out photometric measurements | 8 |
| 1.3.1 | The analysis specifications | 8 |
| 1.3.2 | Reagent blank values | 9 |
| 1.3.3 | Dosing of sample and reagents | 10 |
| 1.4 | Working with the analysis timer | 12 |
| 1.5 | Sample dilution | 12 |
| 1.6 | Minimizing interfering effects | 12 |
| 1.6.1 | General information | 12 |
| 1.6.2 | Influence of the pH value | 13 |
| 1.6.3 | Influence of turbidity | 14 |
| 1.6.4 | Influence of complexing agents | 14 |
| 1.7 | Ordering information | 15 |
| 2 | Analysis specifications | 17 |
| | List of available photometric test sets | 17 |
| | Acid capacity up to pH 4.3 | 21 |
| | Acid capacity up to pH 4.3 | 22 |
| | Aluminium | 23 |
| | Aluminium | 24 |
| | Ammonia | 25 |
| | Ammonia | 26 |
| | Ammonia | 27 |
| | Ammonium | 28 |
| | Ammonium | 29 |
| | Ammonium | 30 |
| | Ammonium | 31 |
| | Ammonium | 32 |
| | Ammonium vario | 33 |
| | Ammonium vario HR | 34 |
| | Ammonium vario LR | 35 |
| | Arsenic | 36 |
| | Cadmium | 37 |
| | Cadmium | 38 |
| | Calcium | 39 |
| | Calcium | 40 |
| | Carbon dioxide | 41 |
| | Carbon dioxide | 42 |
| | Chloride | 43 |
| | Chloride | 44 |
| | Chloride | 45 |

| | |
|------------------------------------|----|
| Chlorine (free & total) | 46 |
| Chlorine (free) | 47 |
| Chlorine (free) vario | 48 |
| Chlorine (free) vario | 49 |
| Chlorine (total) vario | 50 |
| Chlorine dioxide | 51 |
| Chlorine dioxide | 52 |
| Chrome | 53 |
| COD | 54 |
| COD | 55 |
| COD | 56 |
| COD | 57 |
| COD | 58 |
| COD | 59 |
| COD (Hg-free) | 60 |
| COD (Hg-free) | 61 |
| COD | 62 |
| COD | 63 |
| COD HR | 64 |
| COD LR | 65 |
| COD MR | 66 |
| Coloration at 435 nm (FB436) | 67 |
| Coloration at 517 nm (FB517) | 68 |
| Coloration at 610 nm (FB610) | 69 |
| Copper | 70 |
| Copper | 71 |
| Copper | 72 |
| Copper vario | 73 |
| Cyanide (free cyanide) | 74 |
| Fluoride | 75 |
| Formaldehyde | 76 |
| Gold | 77 |
| Iron | 78 |
| Iron | 79 |
| Iron | 80 |
| Iron | 81 |
| Iron vario | 82 |
| Iron vario TPTZ | 83 |
| Lead | 84 |
| Lead | 85 |
| Magnesium | 86 |
| Manganese | 87 |

| | |
|----------------------------------|-----|
| Manganese | 88 |
| Manganese | 89 |
| Manganese vario | 90 |
| Molybdenum | 91 |
| Molybdate vario | 92 |
| Nickel | 93 |
| Nickel | 94 |
| Nitrate | 95 |
| Nitrate | 96 |
| Nitrate | 97 |
| Nitrate | 98 |
| Nitrite | 99 |
| Nitrite | 100 |
| Nitrite | 101 |
| Nitrite | 102 |
| Nitrite | 103 |
| Nitrite HR | 104 |
| Nitrite LR | 105 |
| Nitrite vario | 106 |
| Nitrogen (total) | 107 |
| Nitrogen, total HR | 108 |
| Nitrogen, total LR | 109 |
| Ozone | 110 |
| Ozone | 111 |
| Phenol | 112 |
| Phosphate | 113 |
| Phosphate | 114 |
| Phosphate | 115 |
| Phosphate: Total P | 116 |
| Phosphate: ortho-P | 117 |
| Phosphate: Total P | 118 |
| Phosphate: ortho-P | 119 |
| Phosphate vario (ortho) | 120 |
| Phosphate, ortho | 121 |
| Phosphate, total | 122 |
| Phosphate: ortho-P | 123 |
| Phosphate: ortho-P | 124 |
| Phosphate: Orthophosphate | 125 |
| Phosphate: Orthophosphate | 126 |
| Phosphate: Total phosphate | 127 |
| Phosphate: Total phosphate | 128 |
| Potassium | 129 |

| | |
|--------------------------------------|-----|
| Potassium | 130 |
| Silica HR vario | 131 |
| Silica HR vario | 132 |
| Silica LR vario | 133 |
| Silicon | 134 |
| Silicon | 135 |
| Silicon | 136 |
| Silver | 137 |
| Sodium | 138 |
| Sulfate | 139 |
| Sulfate vario | 140 |
| Tensides (anionic) | 141 |
| Tensides (nonionic) | 142 |
| Water hardness, total hardness | 143 |
| Zinc | 144 |
| Zinc | 145 |

1 Practical part

1.1 For your safety

When developing test sets WTW carefully sees that the tests can be carried out as safely as possible. Some hazards by dangerous substances, however, cannot always be avoided.



Caution

Improper handling of certain reagents can cause damage to your health.

In any case follow the safety labels on the packing and the safety instructions of the package insert.

Protective measures specified there have to be followed exactly.

Qualification of the user

We assume that, due to their professional training and experience, the users are able to correctly understand the safety labels and safety instructions and to appropriately follow the protective measures specified there.

1.2 General information on test sets

Test sets contain special reagents that are added to the test sample according to a certain specification (analysis specification). These reagents react with the test sample. After the reaction the prepared product is inserted in the cell shaft of the photometer in a cell and the photometric measurement is carried out.

In addition to the reagents in the test set, further common laboratory auxiliary reagents can be required such as acids or lyes to adjust a certain pH value.

Categories

The test sets can be divided into two categories depending on the way they are carried out:

- Reaction cell tests

They provide highest convenience. All the special reagents required are prepared in a measuring cell. Normally, a certain quantity of sample has just to be added (with some reaction cell tests, a dose of reagent as well).

- Reagent tests

They contain all reagents required for measurement. Mostly, the sample and reagents can be prepared directly in the cell. An empty cell is required to carry out a reagent test (diameter, see analysis specification).

1.3 Carrying out photometric measurements

1.3.1 The analysis specifications








Following the present practical part are the analysis specifications for all photometric measurements that can be carried out with this photometer.

Program number

In order to measure, you have to enter on the photometer the program number quoted in the analysis specification. If you have connected a bar code reader, you can read in the program number from the analysis specification with it.

Program number

Program number as bar code

| | | | | | | | | | | | |
|---|-------------------------|--|---|-----------------------|-------|------------------|-------------------------|--------------|-------|-------------------------|--------------------|
| COD Program no. | 82 |  |  | | | | | | | | |
| <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 30%;">WTW model no.:</td> <td>C4/25</td> </tr> <tr> <td>Category:</td> <td>KT (reaction cell test)</td> </tr> <tr> <td>Cell:</td> <td>16 mm</td> </tr> <tr> <td>Measuring range:</td> <td>25 - 1500 mg/l COD</td> </tr> </table> <p>Note: Before using the test with your photometer for the first time, determine the reagent blank value.</p> <div style="display: flex; justify-content: space-around; align-items: flex-end;"> <div style="text-align: center;">  <p>Shake the reaction cell so that sediment is</p> </div> <div style="text-align: center;">  <p>Carefully pipette 3.0 ml of sample into the cell.</p> </div> <div style="text-align: center;">  <p>Heat the cell in the thermoreactor for two</p> </div> <div style="text-align: center;">  <p>Remove the cell from the thermoreactor and let it</p> </div> <div style="text-align: center;">  <p>After approx. 10 min cooling time sway the</p> </div> </div> | | | | WTW model no.: | C4/25 | Category: | KT (reaction cell test) | Cell: | 16 mm | Measuring range: | 25 - 1500 mg/l COD |
| WTW model no.: | C4/25 | | | | | | | | | | |
| Category: | KT (reaction cell test) | | | | | | | | | | |
| Cell: | 16 mm | | | | | | | | | | |
| Measuring range: | 25 - 1500 mg/l COD | | | | | | | | | | |

Further specifications

- WTW model number
- Category
- Cell to be used. Use suitable cells from the WTW product range only
- Measuring ranges and citation forms. After measurement you can switch between the stated citation forms.



Note

The measuring ranges specified in this Analysis Manual are valid especially for measurements with the pHotoFlex / pHotoFlex Turb and can be different from other data, e. g. on the package insert.

1.3.2 Reagent blank values

The evaluation of the photometric measurement always refers to the comparison value of a sample without the substance to be determined (reagent blank value). Thus the influence of the basic absorbance of the reagents on photometric measurement is compensated for.

In practice, measurement of the reagent blank value is carried out with the same amount of deionized water instead of sample.

Default reagent blank values

For most tests, the reagent blank value is a constant value. It was determined in the factory and stored in the photometer. You can, however, measure the reagent blank value yourself. The default reagent blank value is then overwritten. The default reagent blank values are restored when the photometer is reset to default settings.

Reagent blank values without factory default

For some tests, it does not make sense to store a default reagent blank value in the factory, e.g. if a change of certain test elements during the storing period cannot be excluded. In this case, a reagent blank value has to be determined before the first measurement with a new photometer. The photometer informs you if no valid reagent blank value is available. The measured reagent blank value remains stored in the meter until a new reagent blank value is determined. When the photometer is reset to default, all reagent blank values that were not stored in the factory are erased.




Note

You can increase accuracy if you determine the reagent blank value with a test of a newly started reagent package and use the reagent blank value for all tests of this package.

1.3.3 Dosing of sample and reagents


The exact dosing of the individual reagents is a precondition for the successful carrying out of a test. The test sets contain simple accessories for correct dosing.

Instructions on using dosing accessories provided with test sets




Dropping bottles (liquid dosing)

Hold the dropping bottle in an exactly vertical position with the dropping tip downward and let the reagent drip out slowly so the correct drop size forms.



Dosing tops/measurers (powder dosing)

Screw the dosing top on the reagent bottle instead of the screw cap. For dosing, hold the reagent bottle in a vertical position and for each specified dose press the lateral slide into the measurer up to the stop once. Subsequently, close the reagent bottle with the original screw cap so the contents cannot become moist.



Microspoon (powder dosing)

Microspoons are integrated in the screw cap of the reagent bottles. They are available in several colors for several dosing quantities.

Precise dosing of liquids


Exact dosing of the sample (and of certain reagents) is especially important. Use suitable laboratory pipettes for this. The two following pipettes from the WTW product range cover the whole range of required dosing volumes:

- Variable piston pipette 100-1000 μl (WTW model KHP/Var 1000)
- Variable piston pipette 0.5-5.0 ml (WTW model KHP/Var 5000)


Read the operating manual of the pipette and make yourself really familiar with the correct dosing procedure. For volumes greater than 5.0 ml, pipette twice if necessary.

Use of powder packs


Some test sets contain reagents readily dosed as powder in small packs. Use the powder packs as follows:




- Hold the powder pack in a vertical position as shown in the picture (perforation up) and hit it slightly against a solid base so the powder collects in the lower part.



- Cut open the pack (tear if necessary) along a horizontal line, starting at the perforation.



- Slightly press the pack from both sides so that a wide opening forms.



- Empty the pack completely. For narrow vessel openings use a suitable funnel top.

1.4 Working with the analysis timer

Many determination procedures include steps that need certain periods of time to expire. All these periods are stored in the photometer. If the analysis timer is switched on, a corresponding timer is activated for each time-critical step according to the proceeding described in the analysis specification. Only after all timers have expired can the photometric measurement be started.

1.5 Sample dilution

Diluting the sample can be required for the following reasons:

- The expected concentration of the substance to be determined is near or above the upper measuring range limit
- Other substances in the sample cause the measured values to be too high or too low due to matrix interferences

With the pHotoFlex (Turb) you can measure diluted samples without having to multiply the measured value by the dilution factor afterwards. The measuring range is extended automatically. To do so, enter the dilution number in the photometer before measuring. Admissible are dilution numbers from 1+1 up to 1+99 (volume parts water + volume parts test sample) in whole number steps.

For diluting, use deionized water and common laboratory dosing equipment with sufficient dosing accuracy (volumetric flasks, pipettes etc.).



Note

Note that the dilution error can increase with an increasing dilution. Therefore, check whether a different test or method with a suitable measuring range can be used instead of diluting the sample.

1.6 Minimizing interfering effects

1.6.1 General information

The following factors can affect photometric determination and cause incorrect measurement results:

- Unsuitable pH value of the sample
- Turbidity
- Interferences due to certain water substances (e.g. complexing agents often disturb the determination of metals)
- Adverse temperature

- Improper, especially not representative, sampling
- Change of the sample due to too long storing period or unsuitable storing conditions until measurement

Many tests have package inserts included. Read these package inserts thoroughly. They inform you of special features to be noted during sampling, preparing the sample and carrying out the test and of possible interferences.

The following chapters provide a detailed description of important influencing factors and practical instructions on remedial actions.

1.6.2 Influence of the pH value

The pH value can affect the course of chemical reactions in a photometric determination. For some tests the pH value of a solution has to be in a certain range. With these tests, the analysis specification informs you of the necessity to check and if necessary adjust the pH value.

Adjusting the pH value

Observe the following points when adjusting the pH range:

- Measure the pH value with the aid of a pH meter or pH indicator
- Use the acids and lyes specified in the analysis specification
- Add the acid or lye drop by drop and measure the pH value after each drop added. Thus the volume of the sample does not increase (is not diluted) too much
- The volume increase by the drops is negligible if the resulting dilution is less than 2 %. With a greater dilution, the measurement result should be converted accordingly. Adding up to five drops per 10 ml of solution is uncritical as a rule of thumb.

1.6.3 Influence of turbidity

With samples that are visibly turbid, measured values can be oscillating or too high or too low during photometrical determination.

Compensating for turbidity

Depending on the type of sample or substance to be determined, the influence of turbidity can be compensated for in different ways:

- If you are positive that the substance to be determined is exclusively in the dissolved part, the sample can be filtrated before carrying out determination. For filtrating, simple common laboratory paper filters or membrane filters (recommended pore size 0.45 µm) can be used.
- If you assume that a considerable part of the substance to be determined is bound in the solid part of the sample, the substance has to be brought into an analysable form before carrying out the photometric determination. This is done in a chemical digestion procedure. Suitable digestion reagents can be found in the WTW catalog. With some procedures, the critical solid substance part is digested during the determination itself (e. g. COD measurement in aqueous samples with suspended matter with parts of organic compounds). Here it is important for an exact determination that the part of suspended matter be representative for the sample. For this the sample has to be homogenized, e.g. with a disperser.

1.6.4 Influence of complexing agents

Complexing agents can disturb the determination of metals by forming very stable compounds with them. In this form the metals cannot be analyzed. Here, a digestion has to be carried out prior to photometric determination. Suitable digestion reagents can be found in the WTW catalog.

1.7 Ordering information

| Description | Model | Order no. |
|--|--------------|-----------|
| 16 mm empty cell (25 pieces) | RK 14/25 | 250 621 |
| Variable piston pipette 100-1000 µl | KHP/Var 1000 | 250 545 |
| Variable piston pipette 0.5-5.0 ml | KHP/Var 5000 | 250 546 |

2 Analysis specifications

List of available photometric test sets

| Name | Prog | Model | Order no. | Measuring range (main citation form) | Cell | Type* | Blank value | P. |
|----------------------------|------|---------------|-----------|---------------------------------------|-------|-------|-------------|----|
| Acid capacity up to pH 4.3 | 29 | 01762 | 252059 | 0.20 - 7.50 mmol/l OH | 16 mm | KT | required | 21 |
| Acid capacity up to pH 4.3 | 105 | 01758 | 252087 | 0.40 - 8.00 mmol/l OH | 16 mm | KT | required | 22 |
| Aluminium | 30 | 14825 | 250425 | 0.05 - 0.40 mg/l Al | 28 mm | RT | required | 23 |
| Aluminium | 323 | Al-1 TP | 251408 | 0.00 - 0.25 mg/l Al | 28 mm | RT | required | 24 |
| Aluminum | 539 | Al-2 | *** | 0,04 - 0,25 mg/l Al | 13 mm | CV | required | ** |
| Ammonia | 71 | 14752 | 250426 | (0.02 - 0.90 mg/l NH ₄ -N) | 28 mm | RT | | 25 |
| Ammonia | 72 | 14752 | 250426 | (0.02 - 1.50 mg/l NH ₄ -N) | 16 mm | RT | | 26 |
| Ammonia | 74 | 14544 | 250329 | (0.5 - 16.0 mg/l NH ₄ -N) | 16 mm | KT | required | 27 |
| Ammonium | 83 | A6/25 | 252072 | 0.20 - 8.00 mg/l NH ₄ -N | 16 mm | KT | required | 28 |
| Ammonium | 31 | 14752 | 250426 | 0.02 - 0.90 mg/l NH ₄ -N | 28 mm | RT | | 29 |
| Ammonium | 32 | 14752 | 250426 | 0.02 - 1.50 mg/l NH ₄ -N | 16 mm | RT | | 30 |
| Ammonium | 48 | 14558 | 252000 | 0.20 - 8.00 mg/l NH ₄ -N | 16 mm | KT | required | 31 |
| Ammonium | 60 | 14544 | 250329 | 0.5 - 16.0 mg/l NH ₄ -N | 16 mm | KT | required | 32 |
| Ammonium | 506 | NH4-4 CV | *** | 0.10 - 3.00 mg/l NH ₄ -N | 13 mm | CV | | ** |
| Ammonium | 507 | NH4-5 CV | *** | 0.50 - 7.00 mg/l NH ₄ -N | 13 mm | CV | | ** |
| Ammonium vario | 324 | NH4-1 TP | 251401 | 0.00 - 0.50 mg/l NH ₄ -N | 28 mm | RT | | 33 |
| Ammonium vario HR | 313 | NH4-3 TC (HR) | 251993 | 0 - 50 mg/l NH ₄ -N | 16 mm | KT | | 34 |
| Ammonium vario LR | 312 | NH4-2 TC (LR) | 251998 | 0.00 - 2.50 mg/l NH ₄ -N | 16 mm | KT | | 35 |
| Arsenic | 75 | 01747 | 252063 | 0.002 - 0.100 mg/l As | 16 mm | RT | | 36 |
| Bromine | 526 | Br2-1 | *** | 0,90 - 9,00 mg/l Br ₂ | 13 mm | CV | | ** |
| Cadmium | 4 | 14834 | 250314 | 0.025 - 1.000 mg/l Cd | 16 mm | KT | | 37 |
| Cadmium | 103 | 01745 | 252051 | 0.010 - 0.500 mg/l Cd | 28 mm | RT | required | 38 |
| Calcium | 62 | 14815 | 250428 | 10 - 160 mg/l Ca | 16 mm | RT | required | 39 |
| Calcium | 63 | 14815 | 250428 | 5 - 80 mg/l Ca | 28 mm | RT | required | 40 |
| Carbon dioxide | 73 | 01762 | 252059 | (0.20 - 7.50 mmol/l OH) | 16 mm | KT | required | 41 |
| Carbon dioxide | 106 | 01758 | 252087 | (0.40 - 8.00 mmol/l OH) | 16 mm | KT | required | 42 |
| Chloride | 104 | 14897 | 250491 | 2.5 - 30.0 mg/l Cl | 16 mm | RT | required | 43 |
| Chloride | 64 | 14897 | 250491 | 10 - 190 mg/l Cl | 16 mm | RT | required | 44 |
| Chloride | 70 | 14730 | 250353 | 5 - 125 mg/l Cl | 16 mm | KT | required | 45 |
| Chlorine (free & total) | 501 | Cl-10 CV | *** | 0,40 - 5.00 mg/l Cl ₂ | 13 mm | CV | | ** |
| Chlorine (free & total) | 34 | 00597 | 250420 | 0.05 - 5.00 mg/l Cl ₂ | 16 mm | KT | | 46 |
| Chlorine (free) | 33 | 00595 | 250419 | 0.05 - 5.00 mg/l Cl ₂ | 16 mm | KT | | 47 |
| Chlorine (free) vario | 325 | Cl-1 TP | 251402 | 0.00 - 2.00 mg/l Cl ₂ | 28 mm | RT | | 48 |
| Chlorine (free) vario | 326 | Cl-2 TP | 251414 | 0.00 - 5.00 mg/l Cl ₂ | 28 mm | RT | | 49 |
| Chlorine (total) vario | 327 | Cl-3 TP | *** | 0.00 - 2.00 mg/l Cl ₂ | 28 mm | RT | | 50 |
| Chlorine dioxide | 38 | 00608 | 252017 | 0.02 - 5.00 mg/l ClO ₂ | 28 mm | RT | | 51 |
| Chlorine dioxide | 39 | 00608 | 252017 | 0.02 - 7.50 mg/l ClO ₂ | 16 mm | RT | | 52 |
| Chlorine dioxide | 502 | ClO21 CV | *** | 0.8 - 11.0 mg/l ClO ₂ | 13 mm | CV | | ** |
| Chromate | 545 | Cr-1 | *** | 0,20 - 3,50 mg/l CrO ₄ | 13 mm | CV | | ** |
| Chromate | 546 | Cr-2 | *** | 0,70 - 13,00 mg/l CrO ₄ | 13 mm | CV | | ** |

| Name | Prog | Model | Order no. | Measuring range (main citation form) | Cell | Type* | Blank value | P. |
|---------------------------------|------|--------------|-----------|--|-------|-------|-------------|----|
| Chrome | 5 | 14552 | 250341 | 0.05 - 2.00 mg/l Cr | 16 mm | KT | | 53 |
| COD | 49 | 14540 | 252001 | 10 - 150 mg/l COD | 16 mm | KT | required | 54 |
| COD | 50 | 14541 | 252002 | 25 - 1500 mg/l COD | 16 mm | KT | required | 55 |
| COD | 96 | 14895 | 250359 | 15 - 300 mg/l COD | 16 mm | KT | required | 56 |
| COD | 97 | 14690 | 250304 | 50 - 500 mg/l COD | 16 mm | KT | required | 57 |
| COD | 98 | 14691 | 250351 | 300 - 3500 mg/l COD | 16 mm | KT | required | 58 |
| COD | 99 | 14555 | 250309 | 500 - 9500 mg/l COD | 16 mm | KT | required | 59 |
| COD | 520 | COD4 CK (LR) | *** | 10 - 150 mg/l COD | 16 mm | KT | required | ** |
| COD | 521 | COD5 CK (LR) | *** | 10 - 150 mg/l COD | 16 mm | KT | required | ** |
| COD | 522 | COD6 CK (MR) | *** | 100 - 1500 mg/l COD | 16 mm | KT | required | ** |
| COD | 523 | COD7 CK (MR) | *** | 100 - 1500 mg/l COD | 16 mm | KT | required | ** |
| COD | 524 | COD8 CK (HR) | *** | 500 - 15000 mg/l COD | 16 mm | KT | required | ** |
| COD | 525 | COD9 CK (HR) | *** | 500 - 15000 mg/l COD | 16 mm | KT | required | ** |
| COD (Hg-free) | 58 | 09772 | 250301 | 10 - 150 mg/l COD | 16 mm | KT | required | 60 |
| COD (Hg-free) | 59 | 09773 | 250306 | 100 - 1500 mg/l COD | 16 mm | KT | required | 61 |
| COD | 81 | C3/25 | 252070 | 10 - 150 mg/l COD | 16 mm | KT | required | 62 |
| COD | 82 | C4/25 | 252071 | 25 - 1500 mg/l COD | 16 mm | KT | required | 63 |
| COD HR | 311 | COD3 TC (HR) | 251997 | 0 - 15000 mg/l COD | 16 mm | KT | required | 64 |
| COD LR | 309 | COD1 TC (LR) | 251991 | 0 - 150 mg/l COD | 16 mm | KT | required | 65 |
| COD MR | 310 | COD2 TC (MR) | 251992 | 0 - 1500 mg/l COD | 16 mm | KT | required | 66 |
| Coloration at 435 nm (FB436) | 43 | FB436 | **** | 0.5 - 50.0 m ⁻¹ | 28 mm | | | 67 |
| Coloration at 517 nm (FB517) | 44 | FB517 | **** | 0.5 - 50.0 m ⁻¹ | 28 mm | | | 68 |
| Coloration at 610 nm (FB610) | 45 | FB610 | **** | 0.5 - 50.0 m ⁻¹ | 28 mm | | | 69 |
| Copper | 13 | 14553 | 250408 | 0.05 - 7.50 mg/l Cu | 16 mm | KT | | 70 |
| Copper | 41 | 14767 | 250441 | 0.04 - 3.50 mg/l Cu | 28 mm | RT | | 71 |
| Copper | 42 | 14767 | 250441 | 0.10 - 6.00 mg/l Cu | 16 mm | RT | | 72 |
| Copper | 505 | Cu-2 CV | *** | 0.5 - 12.0 mg/l Cu | 13 mm | CV | | ** |
| Copper vario | 302 | Cu-1 TP | 251406 | 0.00 - 5.00 mg/l Cu | 28 mm | RT | | 73 |
| Cyanide (free cyanide) | 6 | 14561 | 250344 | 0.01 - 0.30 mg/l CN | 16 mm | KT | | 74 |
| DEHA (N,N-Diethylhydroxylamine) | 528 | DEHA1 | *** | 0,15 - 2,00 mg/l DEHA | 13 mm | CV | | ** |
| Fluoride | 12 | 14557 | 250365 | 0.04 - 1.00 mg/l F | 16 mm | KT | required | 75 |
| Fluoride | 544 | F-1 | *** | 0,30 - 2,00 mg/l F | 13 mm | CV | required | ** |
| Formaldehyde | 92 | 14500 | 250406 | 0.10 - 7.00 mg/l HCHO | 16 mm | KT | required | 76 |
| Glycol | 529 | Gly-1 | *** | 0,60-10,00 mg/l C ₂ H ₆ O ₂ | 13 mm | CV | | ** |
| Gold | 77 | 14821 | 250436 | 0.5 - 9.0 mg/l Au | 16 mm | RT | | 77 |
| Hydrazine | 530 | Hyd-1 | *** | 0,07 - 1,20 mg/l N ₂ H ₄ | 13 mm | CV | | ** |
| Iron | 9 | 14549 | 250349 | 0.05 - 3.00 mg/l Fe | 16 mm | KT | | 78 |
| Iron | 10 | 14761 | 250435 | 0.05 - 1.50 mg/l Fe | 28 mm | RT | | 79 |
| Iron | 11 | 14761 | 250435 | 0.10 - 3.00 mg/l Fe | 16 mm | RT | | 80 |
| Iron | 107 | 14896 | 250361 | 1,0 - 50,0 mg/l Fe | 16 mm | KT | required | 81 |
| Iron (soluble und total) | 504 | Fe-3 CV | *** | 0.2 - 6.0 mg/l Fe | 13 mm | CV | | ** |
| Iron (soluble und total) | 509 | Fe-4 CV | *** | 1.0 - 25.0 mg/l Fe | 13 mm | CV | | ** |
| Iron vario | 301 | Fe-2 TP | 251403 | 0.00 - 3.00 mg/l Fe | 28 mm | RT | | 82 |
| Iron vario TPTZ | 300 | Fe-1 TP | 251405 | 0.00 - 1.80 mg/l Fe | 28 mm | RT | | 83 |
| Lead | 2 | 09717 | 252034 | 0.01 - 4.00 mg/l Pb | 28 mm | RT | required | 84 |

| Name | Prog | Model | Order no. | Measuring range (main citation form) | Cell | Type* | Blank value | P. |
|--------------------|------|--------------------|------------------|---|-------|-------|-------------|-----|
| Lead | 3 | 09717 | 252034 | 0.02 - 5.00 mg/l Pb | 16 mm | RT | required | 85 |
| Magnesium | 47 | 00815 | 252043 | 5.0 - 75.0 mg/l Mg | 16 mm | KT | required | 86 |
| Manganese | 14 | 00816 | 252035 | 0.10 - 5.00 mg/l Mn | 16 mm | KT | | 87 |
| Manganese | 15 | 14770 | 250442 | 0.02 - 5.00 mg/l Mn | 28 mm | RT | | 88 |
| Manganese | 16 | 14770 | 250442 | 0.04 - 9.00 mg/l Mn | 16 mm | RT | | 89 |
| Manganese | 527 | Mn-2 | *** | 2,0 - 30,0 mg/l Mn | 13 mm | CV | | ** |
| Manganese vario | 303 | Mn-1 TP | 251406 | 0.0 - 20.0 mg/l Mn | 28 mm | RT | | 90 |
| Molybdenum | 80 | 00860 | 252040 | 0.02 - 1.00 mg/l Mo | 16 mm | KT | required | 91 |
| Molybdate | 531 | Mo-2 | *** | 1,0 - 25,0 mg/l Mo | 13 mm | CV | | ** |
| Molybdate vario | 304 | Mo-1 TP | 251407 | 0.0 - 35.0 mg/l Mn | 28 mm | RT | | 92 |
| Nickel | 93 | 14554 | 250409 | 0.10 - 6.00 mg/l Ni | 16 mm | KT | required | 93 |
| Nickel | 95 | 14785 | 250443 | 0.10 - 3.80 mg/l Ni | 28 mm | RT | required | 94 |
| Nitrate | 17 | 14542 | 250410 | 0.5 - 14.5 mg/l NO ₃ -N | 16 mm | KT | | 95 |
| Nitrate | 61 | 14556 | 250411 | 0.10 - 2.70 mg/l NO ₃ -N | 16 mm | KT | required | 96 |
| Nitrate | 69 | 14942 | 250422 | 0.2 - 13.0 mg/l NO ₃ -N | 16 mm | KT | required | 97 |
| Nitrate | 314 | NO3-1 TC | 251989 | 0.0 - 30.0 mg/l NO ₃ -N | 16 mm | KT | required | 98 |
| Nitrate | 513 | NO3-2 CV | *** | 0.20 - 1.50 mg/l NO ₃ -N | 13 mm | CV | | ** |
| Nitrate | 514 | NO3-3 CV | *** | 0.20 - 3.00 mg/l NO ₃ -N | 13 mm | CV | | ** |
| Nitrate | 515 | NO3-4 CV | *** | 5 - 60 mg/l NO ₃ | 13 mm | CV | | ** |
| Nitrite | 18 | 14776/1 14776/2 | 250445 250440 | 0.01 - 0.30 mg/l NO ₂ -N | 28 mm | RT | | 99 |
| Nitrite | 19 | 14776/1 14776/2 | 250445 250440 | 0.02 - 0.50 mg/l NO ₂ -N | 16 mm | RT | | 100 |
| Nitrite | 20 | N4/25 | 250343 | 0.02 - 0.55 mg/l NO ₂ -N | 16 mm | KT | | 101 |
| Nitrite | 85 | N5/25 | 252074 | 0.020 - 0.550 mg/l NO ₂ -N | 16 mm | KT | | 102 |
| Nitrite | 55 | 14547 | 252004 | 0.020 - 0.550 mg/l NO ₂ -N | 16 mm | KT | | 103 |
| Nitrite | 516 | NO2-3 CV | *** | 0.08 - 0.80 mg/l NO ₂ -N | 13 mm | CV | | ** |
| Nitrite HR | 317 | NO2-2 TC | 251994 | 0.30 - 3.00 mg/l NO ₂ -N | 16 mm | KT | | 104 |
| Nitrite LR | 318 | NO2-2 TC | 251995 | 0.03 - 0.60 mg/l NO ₂ -N | 16 mm | KT | | 105 |
| Nitrite vario | 305 | NO2-1 TP | 251410 | 0.00 - 0.33 mg/l NO ₂ -N | 28 mm | RT | | 106 |
| Nitrogen (total) | 35 | 14537 | 250358 | 0.50 - 15.00 mg/l N | 16 mm | KT | required | 107 |
| Nitrogen, total HR | 320 | Ntot2 TC (HR) | 251411 | 10 - 150 mg/l N | 16 mm | KT | required | 108 |
| Nitrogen, total LR | 319 | Ntot1 TC (LR) | 251996 | 0.0 - 25.0 mg/l N | 16 mm | KT | required | 109 |
| Oxygen | 532 | O2-2 | *** | 0,20 - 2,00 mg/l O ₂ | 13 mm | CV | | ** |
| Oxygen | 533 | O2-3 | *** | 0,10 - 1,40 mg/l O ₂ | 13 mm | CV | | ** |
| Oxygen | 540 | O2-1 | *** | 2,0 - 15,0 mg/l O ₂ | 13 mm | CV | | ** |
| Ozone | 36 | 00607 | 252016 | 0.01 - 1.80 mg/l O ₃ | 28 mm | RT | | 110 |
| Ozone | 37 | 00607 | 252016 | 0.01 - 3.50 mg/l O ₃ | 16 mm | RT | | 111 |
| Phenol | 91 | 14551 | 250412 | 0.10 - 2.50 mg/l C ₆ H ₅ OH | 16 mm | KT | required | 112 |
| Phenol | 534 | Phen1 | *** | 0,40 - 8,00 mg/l C ₆ H ₅ OH | 13 mm | CV | | ** |
| Phenol | 535 | Phen2 | *** | 1,0 - 20,0 mg/l C ₆ H ₅ OH | 13 mm | CV | | ** |
| Phosphate | 21 | 14546 | 250413 | 0.5 - 25.0 mg/l PO ₄ -P | 16 mm | KT | required | 113 |
| Phosphate | 22 | 14848 | 250446 | 0.20 - 1.50 mg/l PO ₄ -P | 28 mm | RT | | 114 |
| Phosphate | 23 | 14848 | 250446 | 0.40 - 2.50 mg/l PO ₄ -P | 16 mm | RT | | 115 |

| Name | Prog | Model | Order no. | Measuring range (main citation form) | Cell | Type* | Blank value | P. |
|--------------------------------|------|---------------|-----------|--------------------------------------|-------|-------|-------------|-----|
| Phosphate | 503 | PO4-5 CV (HR) | *** | 5 - 40 mg/l PO ₄ | 13 mm | CV | | ** |
| Phosphate | 510 | PO4-4 CV (LR) | *** | 0.20 - 8.00 mg/l PO ₄ | 13 mm | CV | | ** |
| Phosphate: Total P | 86 | P6/25 | 252075 | 0.05 - 3.00 mg/l PO ₄ -P | 16 mm | KT | | 116 |
| Phosphate: ortho-P | 86 | P6/25 | 252075 | 0.05 - 3.00 mg/l PO ₄ -P | 16 mm | KT | | 117 |
| Phosphate: Total P | 87 | P7/25 | 252076 | 0.5 - 15.0 mg/l PO ₄ -P | 16 mm | KT | | 118 |
| Phosphate: ortho-P | 87 | P7/25 | 252076 | 0.5 - 15.0 mg/l PO ₄ -P | 16 mm | KT | | 119 |
| Phosphate vario (ortho) | 306 | PO4-1 TP | 251412 | 0.00 - 0.80 mg/l PO ₄ -P | 28 mm | RT | | 120 |
| Phosphate, ortho | 315 | PO4-2 TC | 251988 | 0.00 - 1.60 mg/l PO ₄ -P | 16 mm | KT | | 121 |
| Phosphate, total | 316 | PO4-3 TC | 251994 | 0.00 - 1.10 mg/l PO ₄ -P | 16 mm | KT | | 122 |
| Phosphate: ortho-P | 78 | 00616 | 252045 | 1.0 - 70.0 mg/l PO ₄ -P | 16 mm | KT | required | 123 |
| Phosphate: ortho-P | 79 | 00798 | 251404 | 1.0 - 50.0 mg/l PO ₄ -P | 16 mm | RT | required | 124 |
| Phosphate: Orthophosphate | 51 | 14543 | 250324 | 0.05 - 3.00 mg/l PO ₄ -P | 16 mm | KT | | 125 |
| Phosphate: Orthophosphate | 53 | 14729 | 250334 | 0.5 - 15.0 mg/l PO ₄ -P | 16 mm | KT | | 126 |
| Phosphate: Total phosphate | 52 | 14543 | 250324 | 0.05 - 3.00 mg/l PO ₄ -P | 16 mm | KT | | 127 |
| Phosphate: Total phosphate | 54 | 14729 | 250334 | 0.5 - 15.0 mg/l PO ₄ -P | 16 mm | KT | | 128 |
| Potassium | 90 | 00615 | 252020 | 30 - 300 mg/l K | 16 mm | KT | required | 129 |
| Potassium | 56 | 14562 | 250407 | 5.00 - 50.00 mg/l K | 16 mm | KT | required | 130 |
| Silica | 536 | Si-3 | *** | 0,30 - 10,00 mg/l Si | 13 mm | CV | | ** |
| Silica HR vario | 307 | Si-2 TP (HR) | 251412 | 0.0 - 70.0 mg/l SiO ₂ | 28 mm | RT | | 131 |
| Silica HR vario | 308 | Si-2 TP (HR) | 251990 | 0 - 100 mg/l SiO ₂ | 16 mm | RT | | 132 |
| Silica LR vario | 321 | Si-1 TP (LR) | 251413 | 0.00 - 1.60 mg/l SiO ₂ | 28 mm | RT | | 133 |
| Silicon | 65 | 14794 | 250438 | 0.10 - 5.00 mg/l Si | 16 mm | RT | | 134 |
| Silicon | 66 | 14794 | 250438 | 0.05 - 2.50 mg/l Si | 28 mm | RT | | 135 |
| Silicon | 67 | 00857 | 252046 | 0.5 - 50.0 mg/l Si | 16 mm | RT | | 136 |
| Silver | 76 | 14831 | 250448 | 0.25 - 2.75 mg/l Ag | 16 mm | RT | required | 137 |
| Sodium | 57 | 00885 | 252044 | 10 - 300 mg/l Na | 16 mm | KT | | 138 |
| Sulfate | 28 | 14548 | 250414 | 25 - 250 mg/l SO ₄ | 16 mm | KT | | 139 |
| Sulfate | 541 | SO4-2 | *** | 8 - 100 mg/l SO ₄ | 13 mm | CV | | ** |
| Sulfate vario | 322 | SO4-1 TP | 251400 | 0 - 70 mg/l SO ₄ | 28 mm | RT | | 140 |
| Sulfide | 542 | S-1 | *** | 0,10 - 3,00 mg/l S | 13 mm | CV | | ** |
| Sulfide | 543 | S-2 | *** | 0,20 - 6,00 mg/l S | 13 mm | CV | | ** |
| Tensides (anionic) | 100 | 14697 | 250333 | 0.05 - 2.00 mg/l MBAS | 16 mm | KT | required | 141 |
| Tensides (nonionic) | 101 | 01787 | 252061 | 0.10 - 7.50 mg/l TritonX-100 | 16 mm | KT | required | 142 |
| Water hardness, total hardness | 46 | 00961 | 252039 | 5 - 215 mg/l GH/Ca | 16 mm | KT | required | 143 |
| Zinc | 40 | 14566 | 250417 | 0.20 - 5.00 mg/l Zn | 16 mm | KT | required | 144 |
| Zinc | 68 | 00861 | 252049 | 0.025 - 1.000 mg/l Zn | 16 mm | KT | required | 145 |
| Zinc | 537 | Zn-1 | *** | 0,30 - 3,00 mg/l Zn | 13 mm | CV | | ** |
| Zinc | 538 | Zn-2 | *** | 0,60 - 6,00 mg/l Zn | 13 mm | CV | | ** |

* KT = reaction cell test; RT = reagent test; CV = ampoule test

** Analysis specification, see package insert

*** Availability of test sets may differ regionally.

More information on test sets is available from your local supplier or from WTW.

**** No test set is required to determine coloration.

Acid capacity up to pH 4.3


Program no.

29




| | |
|------------------|---------------------------------|
| WTW model no.: | 01762 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.20 - 7.50 mmol/l OH |
| | 10 - 375 mg/l CaCO ₃ |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Pipette 5.0 ml of **AC-1** into the cell.



Add 1.0 ml sample with a pipette, close the cell with the screw cap and mix.



Add 0.20 ml **AC-2** with a pipette, close the cell with the screw cap and mix.



Insert the cell in the photometer cell shaft and start measurement.

Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.

Acid capacity up to pH 4.3


Program no.

105




| | |
|------------------|---------------------------------|
| WTW model no.: | 01758 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.40 - 8.00 mmol/l OH |
| | 20 - 400 mg/l CaCO ₃ |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Pipette 4.0 ml of **AC-1** into the cell.



Add 1.0 ml sample with a pipette, close the cell with the screw cap and mix.



Add 0.50 ml **AC-2** with a pipette, close the cell with the screw cap and mix.



Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 14825 |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.05 - 0.40 mg/l Al |
| | Display in mmol/l possible |

Note:


Before using the test with your photometer for the first time, determine the reagent blank value.



Check the pH value of the sample.
Desired range: pH 3-10.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 10 ml sample into the empty cell.




Add 2 level blue microspoons of **Al-1** and dissolve solids.




Add 2.4 ml **Al-2** with a pipette and mix.



Add 0.50 ml **Al-3** with a pipette, close the cell with the screw cap and mix.



Allow to react for 2 minutes.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value for each test set package started.
- For further notes please refer to the package insert of the test.




| | |
|------------------|---------------------|
| WTW model no.: | Al-1 TP |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.00 - 0.25 mg/l Al |

Note: Before using this test with your photometer for the first time, determine the reagent blank value (see notes below).




Pipette 20.0ml sample into the empty cell.



Add the contents of a **VARIO Aluminum ECR F20** powder pack and close the cell with the screw cap.



Dissolve the powder by shaking.




Allow to react for 30 seconds.




Add the contents of a **VARIO Hexamine F20** powder pack and close the cell with the screw cap.




Dissolve the powder by shaking.



For reagent blank value only:
Add 2 drops of **VARIO ECR-Masking RGT**, close the cell with the screw cap and mix.



Allow to react for 5 minutes.



Insert the cell in the photometer cell shaft and start measurement.

Notes:

- To avoid errors due to contamination, rinse the equipment with hydrochloric acid solution (approx. 20%), then deionized water beforehand.
- Adding the **VARIO ECR-Masking RGT** masking reagent is only required when determining the reagent blank value.
- The sample temperature has to be between 20 and 25 °C.
- We recommend to determine a new reagent blank value for each test set package started.



| | |
|-------------------------|---|
| WTW model no.: | 14752 |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | Corresponding to 0.03 - 1.16 mg/l NH_4 or 0.02 - 0.90 mg/l $\text{NH}_4\text{-N}$ |
| | Measuring ranges for NH_3 or $\text{NH}_3\text{-N}$ depending on pH value and temperature, Example: 0.005 - 0.168 mg/l NH_3 at pH 8.5 and 25 °C. |
| | Display in mmol/l possible |

Step 1: pH and temperature measurement

| |
|--------------------------|
| pH & ORP \ pH |
| 10.47 pH |
| 18.6 °C |
| [AR] |

Measuring mode, *pH & ORP*.
Measure pH value and temperature immediately after sampling.

| |
|-------------------|
| Store (927 free) |
| 30.01.06 10:01:58 |
| 10.47 (pH) |
| 18.6 °C |
| assign ID 0 |
| Store (ID 0) |

Store measured values with **<STO>**. If necessary, assign an ID for easier retrieving.

| |
|--|
| Photometry \ Concentration |
| Select program with <PROG> or with Δ ∇ |
| 71: 14752 MR |
| 28 mm NH_3 mg/l |

Switch to the *Photometry* measuring mode and select program no. 71.

| |
|-------------------------------|
| Assign pH and temp. |
| 71: 14752 MR |
| ID = 0 |
| pH = 10.47 |
| Temperature = 18.6 °C |
| Scroll with Δ ∇ |
| Accept |

When the prompt *Assign pH und temp.* appears, select and accept the stored values from the pH and temperature measurement.

| |
|--------------------------------|
| Photometry \ Concentration |
| Insert sample |
| Start measurement with <START> |
| 71: 14752 MR NH_3 |
| 28 mm 0.022 - 1.000 mg/l |

The meter is ready for the photometric measurement (step 2). The measuring range is shown on the display.

Step 2: photometric measurement




Pipette 10.0 ml of sample into the empty cell.




Add 1.20 ml of $\text{NH}_4\text{-1}$ with a pipette and mix.




Add 2 level blue microspoons of $\text{NH}_4\text{-2}$ and close the cell with the screw cap.




Shake the cell vigorously to dissolve solids.




Allow to react for 5 minutes.



Add 8 drops of $\text{NH}_4\text{-3}$, close the cell with the screw cap and mix.



Allow to react for 5 minutes.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- The measuring range depends largely on the pH and temperature. On the basis of the pH and temperature measurement, it is individually calculated and displayed for each determination.
- We recommend to determine a new reagent blank value for each test set package started. To determine the reagent blank value it is not necessary to measure the pH and temperature.
- For further notes please refer to the package insert of the test.




| | |
|-------------------------|---|
| WTW model no.: | 14752 |
| Category: | RT (reagent test) |
| Cell: | 16 mm |
| Measuring range: | Corresponding to 0.03 - 1.93 mg/l NH ₄ or 0.02 - 1.50 mg/l NH ₄ -N |
| | Measuring ranges for NH ₃ or NH ₃ -N depending on pH value and temperature, Example: 0.005 - 0.270 mg/l NH ₃ at pH 8.5 and 25 °C. |
| | Display in mmol/l possible |


Step 1: pH and temperature measurement




Measuring mode, *pH & ORP*.
Measure pH value and temperature immediately after sampling.




Store measured values with **<STO>**. If necessary, assign an ID for easier retrieving.



Switch to the *Photometry* measuring mode and select program no. 72.



When the prompt *Assign pH und temp.* appears, select and accept the stored values from the pH and temperature measurement.




The meter is ready for the photometric measurement (step 2). The measuring range is shown on the display.


Step 2: photometric measurement




Pipette 5.0 ml of sample into the empty cell.




Add 0.60 ml of **NH₄-1** with a pipette and mix.




Add 1 level blue microspoon of **NH₄-2** and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.




Allow to react for 5 minutes.



Add 4 drops of **NH₄-3**, close the cell with the screw cap and mix.



Allow to react for 5 minutes.



Insert the cell in the photometer cell shaft and start measurement.

Notes:

- The measuring range depends largely on the pH and temperature. On the basis of the pH and temperature measurement, it is individually calculated and displayed for each determination.
- We recommend to determine a new reagent blank value for each test set package started. To determine the reagent blank value it is not necessary to measure the pH and temperature.
- For further notes please refer to the package insert of the test.



| | |
|------------------|---|
| WTW model no.: | 14544 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | Corresponding to 0.7 - 20.6 mg/l NH_4 or 0.5 - 16.0 mg/l $\text{NH}_4\text{-N}$ |
| | Measuring ranges for NH_3 or $\text{NH}_3\text{-N}$ depending on pH value and temperature, Example: 0.09 - 3.00 mg/l NH_3 at pH 8.5 and 25 °C. |
| | Display in mmol/l possible |

Note: Before using the test with your photometer for the first time, determine the reagent blank value. To determine the reagent blank value it is not necessary to measure the pH and temperature.

Step 1: pH and temperature measurement

| |
|---------------|
| pH & ORP \ pH |
| 10.47 |
| 18.6 °C |
| [AR] |

| |
|-------------------|
| Store (927 free) |
| 30.01.06 10:01:58 |
| 10.47 (pH) |
| 18.6 °C |
| assign ID |
| Store (ID 0) |

| |
|----------------------------|
| Photometry \ Concentration |
| Select program |
| with <PROG> |
| or with ▲ ▼ |
| 74: 14544 MC |
| 16 mm |
| NH_3 mg/l |

| |
|-----------------------|
| Assign pH and temp. |
| 74: 14544 MC |
| ID = 0 |
| pH = 10.47 |
| Temperature = 18.6 °C |
| Scroll with ▲ ▼ |
| Accept |

| |
|----------------------------|
| Photometry \ Concentration |
| Insert sample |
| Start measurement |
| with <START> |
| 74: 14544 MC |
| 16 mm |
| 0.6 - 17.8 mg/l |
| NH_3 |

Measuring mode, pH & ORP:
Measure pH value and temperature immediately after sampling.


Store measured values with <STO>. If necessary, assign an ID for easier retrieving.

Switch to the *Photometry* measuring mode and select program no. 74.


When the prompt *Assign pH und temp.* appears, select and accept the stored values from the pH and temperature measurement.

The meter is ready for the photometric measurement (step 2). The measuring range is shown on the display.


Step 2: photometric measurement




Pipette 0.50 ml of sample into a reaction cell and mix.




Add 1 dose of $\text{NH}_4\text{-1K}$ with the blue measurer and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 15 minutes.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- The measuring range depends largely on the pH and temperature. On the basis of the pH and temperature measurement, it is individually calculated and displayed for each determination.
- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- The test solution should be yellowish green or green. Very high ammonium concentrations in the sample cause turquoise solutions and too low measured values. Dilute the sample in this case.
- For further notes please refer to the package insert of the test.




| | |
|------------------|---|
| WTW model no.: | A6/25 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.20 - 8.00 mg/l $\text{NH}_4\text{-N}$ |
| | 0.26 - 10.30 mg/l NH_4 |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.



Pipette 1.0 ml of sample into a reaction cell and mix.




Add 1 dose of $\text{NH}_4\text{-1K}$ with the blue measurer and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 15 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|-------------------------|---|
| WTW model no.: | 14752 |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.02 - 0.90 mg/l $\text{NH}_4\text{-N}$ |
| | 0.03 - 1.16 mg/l NH_4 |
| | Display in mmol/l possible |




Pipette 10.0 ml of sample into the empty cell.




Add 1.20 ml of $\text{NH}_4\text{-1}$ with a pipette and mix.




Add 2 level blue microspoons of $\text{NH}_4\text{-2}$ and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.




Allow to react for 5 minutes.



Add 8 drops of $\text{NH}_4\text{-3}$, close the cell with the screw cap and mix.



Allow to react for 5 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value for each test set package started.
- For further notes please refer to the package insert of the test.




| | |
|------------------|---|
| WTW model no.: | 14752 |
| Category: | RT (reagent test) |
| Cell: | 16 mm |
| Measuring range: | 0.02 - 1.50 mg/l $\text{NH}_4\text{-N}$ |
| | 0.03 - 1.93 mg/l NH_4 |
| | Display in mmol/l possible |




Pipette 5.0 ml of sample into the empty cell.




Add 0.60 ml of $\text{NH}_4\text{-1}$ with a pipette and mix.




Add 1 level blue microspoon of $\text{NH}_4\text{-2}$ and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.




Allow to react for 5 minutes.



Add 4 drops of $\text{NH}_4\text{-3}$, close the cell with the screw cap and mix.



Allow to react for 5 minutes.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value for each test set package started.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-------------------------------------|
| WTW model no.: | 14558 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.20 - 8.00 mg/l NH ₄ -N |
| | 0.26 - 10.30 mg/l NH ₄ |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Pipette 1.0 ml of sample into a reaction cell and mix.




Add 1 dose of **NH₄-1K** with the blue measurer and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 15 minutes.



Insert the cell in the photometer cell shaft and start measurement.

Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|------------------------------------|
| WTW model no.: | 14544 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.5 - 16.0 mg/l NH ₄ -N |
| | 0.7 - 20.6 mg/l NH ₄ |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.



Pipette 0.50 ml of sample into a reaction cell and mix.




Add 1 dose of **NH₄-1K** with the blue measurer and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 15 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- The test solution should be yellowish green or green. Very high ammonium concentrations in the sample cause turquoise solutions and too low measured values. Dilute the sample in this case.
- For further notes please refer to the package insert of the test.




| | |
|-------------------------|-------------------------------------|
| WTW model no.: | NH4-1 TP |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.00 - 0.50 mg/l NH ₄ -N |
| | 0.00 - 0.64 mg/l NH ₄ |
| | Display in mmol/l possible |



Pipette 10.0 ml of sample into the empty cell.




Add the contents of a **VARIO AMMONIA Salicylate F10** powder pack and close the cell with the screw cap.



Allow to react for 3 minutes (reaction time).




Add the contents of a **VARIO AMMONIA Cyanurate F10** powder pack and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 15 minutes (reaction time).



Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- If NH₄-N is present in the sample, the solution becomes green after VARIO AMMONIA Cyanurate F10 was added.
- If chlorine is present, sodium thiosulfate has to be added to the sample immediately after sampling. Add 1 drop of a 0.1 mol/l sodium thiosulfate solution per 0.3 mg/l chlorine to 1 liter sample.




| | |
|-------------------------|---------------------------------|
| WTW model no.: | NH4-3 TC (HR) |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0 - 50 mg/l NH ₄ -N |
| | 0.0 - 64.4 mg/l NH ₄ |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired value: approx. pH 7.
Correct with diluted sodium hydroxide solution or hydrochloric acid as necessary.




Pipette 0.1 ml of sample into a reaction cell.



Add the contents of a **VARIO AMMONIA Salicylate F5** powder pack.




Add the contents of a **VARIO AMMONIA Cyanurate F5** powder pack and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 20 minutes.



Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- If NH₄-N is present in the sample, the solution becomes green after VARIO AMMONIA Cyanurate F5 was added.
- If chlorine is present, sodium thiosulfate has to be added to the sample immediately after sampling. Add 1 drop of a 0.1 mol/l sodium thiosulfate solution per 0.3 mg/l chlorine to 1 liter sample.
- Iron disturbs the measurement and can be eliminated as follows: Determine the total iron concentration and prepare an iron standard solution with the determined concentration. Use this solution instead of distilled water to determine the reagent blank value for ammonium measurement.




| | |
|-------------------------|-------------------------------------|
| WTW model no.: | NH4-2 TC (LR) |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.00 - 2.50 mg/l NH ₄ -N |
| | 0.00 - 3.22 mg/l NH ₄ |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired value: approx. pH 7.
Correct with diluted sodium hydroxide solution or hydrochloric acid as necessary.




Pipette 2.0 ml of sample into a reaction cell.



Add the contents of a **VARIO AMMONIA Salicylate F5** powder pack.




Add the contents of a **VARIO AMMONIA Cyanurate F5** powder pack and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 20 minutes.



Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- If NH₄-N is present in the sample, the solution becomes green after VARIO AMMONIA Cyanurate F5 was added.
- If chlorine is present, sodium thiosulfate has to be added to the sample immediately after sampling. Add 1 drop of a 0.1 mol/l sodium thiosulfate solution per 0.3 mg/l chlorine to 1 liter sample.
- Iron disturbs the measurement and can be eliminated as follows: Determine the total iron concentration and prepare an iron standard solution with the determined concentration. Use this solution instead of distilled water to determine the reagent blank value for ammonium measurement.



| | |
|-------------------------|----------------------------|
| WTW model no.: | 01747 |
| Category: | RT (reagent test) |
| Cell: | 16 mm |
| Measuring range: | 0.002 - 0.100 mg/l As |
| | Display in mmol/l possible |



Fill 350 ml of the sample into an Erlenmeyer flask with ground joint.




Add 5 drops of **As-1** and mix.




Add 20 ml sulfuric acid (95-97 % for analysis) and mix.




Add one level red measuring spoon of **As-3** and dissolve solids.



Add 1.0 ml **As-4** with a pipette and mix.




Pipette 5.0 ml **As-5** into the AS absorption tube (WTW article no. 252 066).




Using a pipette, add 1.0 ml **As-6** to the solution in the Erlenmeyer flask and mix.




Add two level red measuring spoons of granulated zinc and **immediately** attach the filled absorption tube on the Erlenmeyer flask.



Allow to react for 2 hours. Sway the solution occasionally or stir slowly with a magnetic stirrer.



Fill the solution from the absorption tube into an empty cell.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value for each test set package started.
- For further notes please refer to the package insert of the test.




| | |
|-------------------------|----------------------------|
| WTW model no.: | 14834 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.025 - 1.000 mg/l Cd |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 3-9.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 5.0 ml of sample into a reaction cell and mix.




Add 0.20 ml **Cd-1K** with a pipette and mix.




Add 1 level green microspoon of **Cd-2K** and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 2 minutes.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|-------------------------|----------------------------|
| WTW model no.: | 01745 |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.010 - 0.500 mg/l Cd |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Check the pH value of the sample.
Required range: pH 3-9.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.



Pipette 1.0 ml of **Cd-1** into the empty cell.




Add 10.0 ml of sample with a pipette and mix.



Add 0.20 ml **Cd-2** with a pipette and mix.




Add 1 level green microspoon of **Cd-3** and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 2 minutes.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 14815 |
| Category: | RT (reagent test) |
| Cell: | 16 mm |
| Measuring range: | 10 - 160 mg/l Ca |
| | Display in mmol/l possible |

Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Check the pH value of the sample.
Desired range: pH 4-10.
Correct with diluted sodium hydroxide solution or hydrochloric acid as necessary.




Pipette 0.10 ml of sample into the empty cell.




Add 5.0 ml **Ca-1** with a pipette and mix.




Add 4 drops of **Ca-2** and mix.



Add 4 drops of **Ca-3** and mix.



Allow to react for **exactly** 8 minutes.
The reaction time has to be kept exactly!



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 14815 |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 5 - 80 mg/l Ca |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Check the pH value of the sample.
Desired range: pH 4-10.
Correct with diluted sodium hydroxide solution or hydrochloric acid as necessary.




Pipette 0.20 ml of sample into the empty cell.




Add 10.0 ml **Ca-1** with a pipette and mix.




Add 8 drops of **Ca-2** and mix.



Add 8 drops of **Ca-3** and mix.



Allow to react for **exactly** 8 minutes.
The reaction time has to be kept exactly!



Insert the cell in the photometer cell shaft and start measurement.

Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.



| | |
|------------------|--|
| WTW model no.: | 01762 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | Corresponding to 0.20 - 7.50 mmol/l OH |
| | Measuring range for CO ₂ depending on pH value and temperature, Example: 7 - 233 mg/l CO ₂ at pH 6.5 and 25 °C. |
| | Display in mmol/l possible |

Note: Before using the test with your photometer for the first time, determine the reagent blank value. To determine the reagent blank value it is not necessary to measure the pH and temperature.

Step 1: pH and temperature measurement

| |
|---------------|
| pH & ORP \ pH |
| 6.50 |
| 18.6 °C |
| [AR] |

| |
|-------------------|
| Store (926 free) |
| 30.01.06 10:13:05 |
| 6.50 (pH) |
| 18.6 °C |
| assign ID |
| Store (ID 0) |

| |
|----------------------------|
| Photometry \ Concentration |
| Select program |
| with <PROG> |
| or with ▲ ▼ |
| 73: 01762 MC |
| CO ₂ |
| 16 mm |
| mg/l |

| |
|-----------------------|
| Assign pH and temp. |
| 73: 01762 MC |
| ID = 0 |
| pH = 6.50 |
| Temperature = 18.6 °C |
| Scroll with ▲ ▼ |
| Accept |

| |
|----------------------------|
| Photometry \ Concentration |
| Insert sample |
| Start measurement |
| with <START> |
| 73: 01762 MC |
| CO ₂ |
| 16 mm |
| 7 - 258 mg/l |

Measuring mode, pH & ORP:
Measure pH value and temperature immediately after sampling.


Store measured values with <STO>. If necessary, assign an ID for easier retrieving.

Switch to the *Photometry* measuring mode and select program no. 73.


When the prompt *Assign pH und temp.* appears, select and accept the stored values from the pH and temperature measurement.

The meter is ready for the photometric measurement (step 2). The measuring range is shown on the display.

Step 2: photometric measurement




Pipette 5.0 ml of **AC-1** into the cell.



Add 1.0 ml sample with a pipette, close the cell with the screw cap and mix.



Add 0.20 ml **AC-2** with a pipette, close the cell with the screw cap and mix.



Insert the cell in the photometer cell shaft and start measurement.

Notes:

- The measuring range depends largely on the pH and temperature. On the basis of the pH and temperature measurement, it is individually calculated and displayed for each determination.
- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.



| | |
|------------------|--|
| WTW model no.: | 01758 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | Corresponding to 0.40 - 8.00 mmol/l OH |
| | Measuring range for CO ₂ depending on pH value and temperature, |
| | Example: 14 - 275 mg/l CO ₂ at pH 6.5 and 18.6 °C |
| | Display in mmol/l possible |

Note: Before using the test with your photometer for the first time, determine the reagent blank value. To determine the reagent blank value it is not necessary to measure the pH and temperature.

Step 1: pH and temperature measurement

| | | | | |
|---|---|---|---|---|
| <p>pH & ORP \ pH</p> <p>6.50</p> <p>18.6 °C</p> <p>[AR]</p> | <p>Store (926 free)</p> <p>30.01.06 10:13:05</p> <p>6.50 (pH)</p> <p>18.6 °C</p> <p>assign ID 0</p> <p>Store (ID 0)</p> | <p>Photometry \ Concentration</p> <p>Select program with <PROG> or with ▲ ▼</p> <p>73: 01762 MC</p> <p>16 mm</p> <p>CO₂ mg/l</p> | <p>Assign pH and temp.</p> <p>73: 01762 MC</p> <p>ID = 0</p> <p>pH = 6.50</p> <p>Temperature = 18.6 °C</p> <p>Scroll with ▲ ▼</p> <p>Accept</p> | <p>Photometry \ Concentration</p> <p>Insert sample</p> <p>Start measurement with <START></p> <p>73: 01762 MC</p> <p>16 mm</p> <p>7 - 258 mg/l</p> <p>CO₂</p> |
|---|---|---|---|---|

Measuring mode, pH & ORP:
Measure pH value and temperature immediately after sampling.


Store measured values with <STO>. If necessary, assign an ID for easier retrieving.

Switch to the *Photometry* measuring mode and select program no. 106.


When the prompt *Assign pH und temp.* appears, select and accept the stored values from the pH and temperature measurement.

The meter is ready for the photometric measurement (step 2). The measuring range is shown on the display.

Step 2: photometric measurement




Pipette 4.0 ml of **AC-1** into the cell.



Add 1.0 ml sample with a pipette, close the cell with the screw cap and mix.



Add 0.50 ml **AC-2** with a pipette, close the cell with the screw cap and mix.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- The measuring range depends largely on the pH and temperature. On the basis of the pH and temperature measurement, it is individually calculated and displayed for each determination.
- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 14897 |
| Category: | RT (reagent test) |
| Cell: | 16 mm |
| Measuring range: | 2.5 - 30.0 mg/l Cl |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Check the pH value of the sample.
Required range:
pH 1-12.
Correct with diluted ammonia solution or nitric acid as necessary.




Pipette 5.0 ml of sample into the empty cell.




Add 2.5 ml **Cl-1** with a pipette and mix.



Add 0.50 ml **Cl-2** with a pipette and mix.



Allow to react for 1 minute.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value for each test set package started.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 14897 |
| Category: | RT (reagent test) |
| Cell: | 16 mm |
| Measuring range: | 10 - 190 mg/l Cl |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Check the pH value of the sample.
Desired range: pH 1-12.
Correct with diluted ammonia solution or nitric acid as necessary.




Pipette 1.0 ml of sample into the empty cell.




Add 2.5 ml **Cl-1** with a pipette and mix.



Add 0.50 ml **Cl-2** with a pipette and mix.



Allow to react for 1 minute.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 14730 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 5 - 125 mg/l Cl |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Check the pH value of the sample.
Desired range: pH 1-12.
Correct with diluted ammonia solution or nitric acid as necessary.



Using a pipette, add 0,50 ml of **CI-1K** into a reaction cell, close the cell with the screw cap and mix.



Add 1.0 ml of sample with a pipette, close the cell with the screw cap and mix.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|--------------------------------|
| WTW model no.: | 00597 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.05 - 5.00 mg/l Cl_2 |
| | Display in mmol/l possible |


Determination of total chlorine:




Pipette 5.0 ml of sample into a reaction cell.




Add 1 level blue microspoon of $\text{Cl}_2\text{-1}$ and close the cell with the screw cap.




Shake the cell vigorously to dissolve solids. A small amount of solids may remain undissolved.



Add 2 drops of $\text{Cl}_2\text{-2}$, close the cell with the screw cap and mix.



Allow to react for 3 minutes.



Insert the cell in the photometer cell shaft and start measurement.

Determination of free chlorine:


Similar preparation to above but without adding $\text{Cl}_2\text{-2}$.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- The test sample should be red. Very high chlorine concentrations in the sample cause yellow solutions and too low measured values. Dilute the sample in this case.
- For further notes please refer to the package insert of the test.




| | |
|------------------|--------------------------------|
| WTW model no.: | 00595 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.05 - 5.00 mg/l Cl_2 |
| | Display in mmol/l possible |




Pipette 5.0 ml of sample into a reaction cell.




Add 1 level blue microspoon of $\text{Cl}_2\text{-1}$ and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids. A small amount of solids may remain undissolved.



Allow to react for 3 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- The test sample should be red. Very high chlorine concentrations in the sample cause yellow solutions and too low measured values. Dilute the sample in this case.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------------|
| WTW model no.: | CI-1 TP |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.00 - 2.00 mg/l Cl ₂ |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 6-7.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 10.0 ml of sample into the empty cell.



Add the contents of a **VARIO Chlorine FREE-DPD/F10** powder pack and close the cell with the screw cap.



Mix the contents by swaying (for 20 seconds).




Insert the cell in the photometer cell shaft within one minute and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- The test sample should be pink. Very high chlorine concentrations in the sample cause yellow solutions and too low measured values. Dilute the sample in this case.




| | |
|------------------|----------------------------------|
| WTW model no.: | Cl-2 TP |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.00 - 5.00 mg/l Cl ₂ |
| | Display in mmol/l possible |



Check the pH value of the sample.
Desired range: pH 6-7.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 10.0 ml of sample into an empty beaker.




Add the contents of a **VARIO Chlorine FREE-DPD F25** powder pack and dissolve them by stirring.



Add 15.0 ml deionized water with a pipette and mix.



Fill an empty cell with the prepared sample and close it with the screw cap.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- The test sample should be pink. Very high chlorine concentrations in the sample cause yellow solutions and too low measured values. Dilute the sample in this case.




| | |
|------------------|----------------------------------|
| WTW model no.: | Cl-3 TP |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.00 - 2.00 mg/l Cl ₂ |
| | Display in mmol/l possible |




Pipette 10.0 ml of sample into the empty cell.




Add the contents of one **VARIO Chlorine Total DPD PP** powder pack and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids. A small amount of solids may remain undissolved.



Allow to react for 3 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- The test sample should be pink. Very high chlorine concentrations in the sample cause yellow solutions and too low measured values. Dilute the sample in this case.




| | |
|------------------|-----------------------------------|
| WTW model no.: | 00608 |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.02 - 5.00 mg/l ClO ₂ |
| | Display in mmol/l possible |



Check the pH value of the sample.
Desired range: pH 4-8.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 10.0 ml sample into the empty cell.




Add 2 drops of ClO₂-1 and mix.




Add one level blue microspoon of ClO₂-2 and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 3 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-----------------------------------|
| WTW model no.: | 00608 |
| Category: | RT (reagent test) |
| Cell: | 16 mm |
| Measuring range: | 0.02 - 7.50 mg/l ClO ₂ |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 4-8.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 10.0 ml sample into the empty cell.




Add 2 drops of ClO₂-1 and mix.




Add one level blue microspoon of ClO₂-2 and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 3 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 14552 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.05 - 2.00 mg/l Cr |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 1-9.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Add 6 drops of **Cr-3K** into a reaction cell and close with screw cap.




Shake the cell vigorously to dissolve solids.




Allow to react for 1 minute.



Add 5.0 ml sample with a pipette, close the cell with the screw cap and mix.



Allow to react for 1 minute.



Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-------------------------|
| WTW model no.: | 14540 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 10 - 150 mg/l COD |

Note:


Before using the test with your photometer for the first time, determine the reagent blank value.



Shake the reaction cell so that sediment is suspended.




Carefully pipette 3.0 ml of sample into the cell, close with screw cap and mix vigorously.
Caution, cell becomes very hot!




Heat the cell in the thermoreactor for two hours at 148 °C.




Remove the cell from the thermoreactor and let it cool down in a cell rack.



After approx. 10 min cooling time sway the cell.



Place the cell in the cell rack again and let it cool down **to room temperature**.



Carefully insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- The chloride content of the sample must not exceed 2000 mg/l.
- Homogenize samples containing suspended matter with a disperser.
- Before being inserted in the thermoreactor and for photometric measurements the outside of the cell must be free of any contamination (e.g. fingerprints or drops of water). Wipe the cell with a dry cloth as necessary.
- Let the cell cool down long enough (at least 45 min) before inserting it in the photometer cell shaft. The cells remain stable for a long time after reaction and can also be left overnight and then measured.
- After cooling do not rock the cell until the measurement takes place in order not to suspend the solids that formed during the reaction. Suspended matter disturbs the photometric measurement.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-------------------------|
| WTW model no.: | 14541 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 25 - 1500 mg/l COD |

Note:


Before using the test with your photometer for the first time, determine the reagent blank value.



Shake the reaction cell so that sediment is suspended.




Carefully pipette 3.0 ml of sample into the cell, close with screw cap and mix vigorously.
Caution, cell becomes very hot!




Heat the cell in the thermoreactor for two hours at 148 °C.




Remove the cell from the thermoreactor and let it cool down in a cell rack.



After approx. 10 min cooling time sway the cell.



Place the cell in the cell rack again and let it cool down **to room temperature**.



Carefully insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- The chloride content of the sample must not exceed 2000 mg/l.
- Homogenize samples containing suspended matter with a disperser.
- Before being inserted in the thermoreactor and for photometric measurements the outside of the cell must be free of any contamination (e.g. fingerprints or drops of water). Wipe the cell with a dry cloth as necessary.
- Let the cell cool down long enough (at least 45 min) before inserting it in the photometer cell shaft. The cells remain stable for a long time after reaction and can also be left overnight and then measured.
- After cooling do not rock the cell until the measurement takes place in order not to suspend the solids that formed during the reaction. Suspended matter disturbs the photometric measurement.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-------------------------|
| WTW model no.: | 14895 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 15 - 300 mg/l COD |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.



Shake the reaction cell so that the sediment is suspended.




Carefully pipette 2.0 ml of sample into the cell, close with screw cap and mix vigorously.
Caution, cell becomes very hot!




Heat the cell in the thermoreactor for two hours at 148 °C.




Remove the cell from the thermoreactor and let it cool down in a cell rack.



After approx. 10 min cooling time sway the cell.



Place the cell in the cell rack again and let it cool down **to room temperature**.



Carefully insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- The chloride content of the sample must not exceed 2000 mg/l.
- Homogenize samples containing suspended matter with a disperser.
- Before being inserted in the thermoreactor and for photometric measurements the outside of the cell must be free of any contamination (e.g. fingerprints or drops of water). Wipe the cell with a dry cloth as necessary.
- Let the cell cool down long enough (at least 45 min) before inserting it in the photometer cell shaft. The cells remain stable for a long time after reaction and can also be left overnight and then measured.
- After cooling do not rock the cell until the measurement takes place in order not to suspend the solids that formed during the reaction. Suspended matter disturbs the photometric measurement.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-------------------------|
| WTW model no.: | 14690 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 50 - 500 mg/l COD |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.



Shake the reaction cell so that the sediment is suspended.




Carefully pipette 2.0 ml of sample into the cell, close with screw cap and mix vigorously.
Caution, cell becomes very hot!




Heat the cell in the thermoreactor for two hours at 148 °C.




Remove the cell from the thermoreactor and let it cool down in a cell rack.



After approx. 10 min cooling time sway the cell.



Place the cell in the cell rack again and let it cool down **to room temperature**.



Carefully insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- The chloride content of the sample must not exceed 2500 mg/l.
- Homogenize samples containing suspended matter with a disperser.
- Before being inserted in the thermoreactor and for photometric measurements the outside of the cell must be free of any contamination (e.g. fingerprints or drops of water). Wipe the cell with a dry cloth as necessary.
- Let the cell cool down long enough (at least 45 min) before inserting it in the photometer cell shaft. The cells remain stable for a long time after reaction and can also be left overnight and then measured.
- After cooling do not rock the cell until the measurement takes place in order not to suspend the solids that formed during the reaction. Suspended matter disturbs the photometric measurement.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-------------------------|
| WTW model no.: | 14691 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 300 - 3500 mg/l COD |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.



Shake the reaction cell so that the sediment is suspended.




Carefully pipette 2.0 ml of sample into the cell, close with screw cap and mix vigorously.
Caution, cell becomes very hot!




Heat the cell in the thermoreactor for two hours at 148 °C.




Remove the cell from the thermoreactor and let it cool down in a cell rack.



After approx. 10 min cooling time sway the cell.



Place the cell in the cell rack again and let it cool down **to room temperature**.



Carefully insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- The chloride content of the sample must not exceed 2500 mg/l.
- Homogenize samples containing suspended matter with a disperser.
- Before being inserted in the thermoreactor and for photometric measurements the outside of the cell must be free of any contamination (e.g. fingerprints or drops of water). Wipe the cell with a dry cloth as necessary.
- Let the cell cool down long enough (at least 45 min) before inserting it in the photometer cell shaft. The cells remain stable for a long time after reaction and can also be left overnight and then measured.
- After cooling do not rock the cell until the measurement takes place in order not to suspend the solids that formed during the reaction. Suspended matter disturbs the photometric measurement.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-------------------------|
| WTW model no.: | 14555 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 500 - 9500 mg/l COD |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.



Shake the reaction cell so that the sediment is suspended.




Carefully pipette 1.0 ml of sample into the cell, close with screw cap and mix vigorously.
Caution, cell becomes very hot!




Heat the cell in the thermoreactor for two hours at 148 °C.




Remove the cell from the thermoreactor and let it cool down in a cell rack.



After approx. 10 min cooling time sway the cell.



Place the cell in the cell rack again and let it cool down **to room temperature**.



Carefully insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- The chloride content of the sample must not exceed 5000 mg/l.
- Homogenize samples containing suspended matter with a disperser.
- Before being inserted in the thermoreactor and for photometric measurements the outside of the cell must be free of any contamination (e.g. fingerprints or drops of water). Wipe the cell with a dry cloth as necessary.
- Let the cell cool down long enough (at least 45 min) before inserting it in the photometer cell shaft. The cells remain stable for a long time after reaction and can also be left overnight and then measured.
- After cooling do not rock the cell until the measurement takes place in order not to suspend the solids that formed during the reaction. Suspended matter disturbs the photometric measurement.
- For further notes please refer to the package insert of the test.



| | |
|------------------|-------------------------|
| WTW model no.: | 09772 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 10 - 150 mg/l COD |


Note:

Before using the test with your photometer for the first time, determine the reagent blank value.



Carefully pipette 2.0 ml of sample into the cell, close with screw cap and mix vigorously.


Caution, cell becomes very hot!



Heat the cell in the thermoreactor for two hours at 148 °C.




Remove the cell from the thermoreactor and let it cool down in a cell rack.



After approx. 10 min cooling time sway the cell.



Place the cell in the cell rack again and let it cool down **to room temperature**.



Carefully insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- The presence of chloride causes the measured value to be too high. Refer to the package insert for details.
- Homogenize samples containing suspended matter with a disperser.
- Before being inserted in the thermoreactor and for photometric measurements the outside of the cell must be free of any contamination (e.g. fingerprints or drops of water). Wipe the cell with a dry cloth as necessary.
- Let the cell cool down long enough (at least 45 min) before inserting it in the photometer cell shaft. The cells remain stable for a long time after reaction and can also be left overnight and then measured.
- After cooling do not rock the cell until the measurement takes place in order not to suspend the solids that formed during the reaction. Suspended matter disturbs the photometric measurement.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-------------------------|
| WTW model no.: | 09773 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 100 - 1500 mg/l COD |

Note:

Before using the test with your photometer for the first time, determine the reagent blank value.




Carefully pipette 2.0 ml of sample into the cell, close with screw cap and mix vigorously.
Caution, cell becomes very hot!



Heat the cell in the thermoreactor for two hours at 148 °C.




Remove the cell from the thermoreactor and let it cool down in a cell rack.



After approx. 10 min cooling time sway the cell.



Place the cell in the cell rack again and let it cool down **to room temperature**.



Carefully insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- The presence of chloride causes the measured value to be too high. Refer to the package insert for details.
- Homogenize samples containing suspended matter with a disperser.
- Before being inserted in the thermoreactor and for photometric measurements the outside of the cell must be free of any contamination (e.g. fingerprints or drops of water). Wipe the cell with a dry cloth as necessary.
- Let the cell cool down long enough (at least 45 min) before inserting it in the photometer cell shaft. The cells remain stable for a long time after reaction and can also be left overnight and then measured.
- After cooling do not rock the cell until the measurement takes place in order not to suspend the solids that formed during the reaction. Suspended matter disturbs the photometric measurement.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-------------------------|
| WTW model no.: | C3/25 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 10 - 150 mg/l COD |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.



Shake the reaction cell so that sediment is suspended.




Carefully pipette 3.0 ml of sample into the cell, close with screw cap and mix vigorously.
Caution, cell becomes very hot!




Heat the cell in the thermoreactor for two hours at 148 °C.




Remove the cell from the thermoreactor and let it cool down in a cell rack.



After approx. 10 min cooling time sway the cell.



Place the cell in the cell rack again and let it cool down **to room temperature**.



Carefully insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- The chloride content of the sample must not exceed 2000 mg/l.
- Homogenize samples containing suspended matter with a disperser.
- Before being inserted in the thermoreactor and for photometric measurements the outside of the cell must be free of any contamination (e.g. fingerprints or drops of water). Wipe the cell with a dry cloth as necessary.
- Let the cell cool down long enough (at least 30 min) before inserting it in the photometer cell shaft. The cells remain stable for a long time after reaction and can also be left overnight and then measured.
- After cooling do not rock the cell until the measurement takes place in order not to suspend the solids that formed during the reaction. Suspended matter disturbs the photometric measurement.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-------------------------|
| WTW model no.: | C4/25 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 25 - 1500 mg/l COD |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.



Shake the reaction cell so that sediment is suspended.




Carefully pipette 3.0 ml of sample into the cell, close with screw cap and mix vigorously.
Caution, cell becomes very hot!




Heat the cell in the thermoreactor for two hours at 148 °C.




Remove the cell from the thermoreactor and let it cool down in a cell rack.



After approx. 10 min cooling time sway the cell.



Place the cell in the cell rack again and let it cool down **to room temperature**.



Carefully insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- The chloride content of the sample must not exceed 2000 mg/l.
- Homogenize samples containing suspended matter with a disperser.
- Before being inserted in the thermoreactor and for photometric measurements the outside of the cell must be free of any contamination (e.g. fingerprints or drops of water). Wipe the cell with a dry cloth as necessary.
- Let the cell cool down long enough (at least 30 min) before inserting it in the photometer cell shaft. The cells remain stable for a long time after reaction and can also be left overnight and then measured.
- After cooling do not rock the cell until the measurement takes place in order not to suspend the solids that formed during the reaction. Suspended matter disturbs the photometric measurement.
- For further notes please refer to the package insert of the test.




| | |
|-------------------------|-------------------------|
| WTW model no.: | COD3 TC (HR) |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0 - 15000 mg/l COD |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.



Shake the reaction cell so that sediment is suspended.




Carefully pipette 0.2 ml of sample into the cell, close with screw cap and mix vigorously.
Caution, cell becomes very hot!




Heat the cell in the thermoreactor for two hours at 148 °C.




Remove the cell from the thermoreactor and let it cool down in a cell rack.



After approx. 10 min cooling time sway the cell.



Place the cell in the cell rack again and let it cool down **to room temperature**.



Carefully insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- The chloride content of the sample must not exceed 10,000 mg/l.
- Homogenize samples containing suspended matter with a disperser.
- Before being inserted in the thermoreactor and for photometric measurements the outside of the cell must be free of any contamination (e.g. fingerprints or drops of water). Wipe the cell with a dry cloth as necessary.
- Let the cell cool down long enough (at least 45 min) before inserting it in the photometer cell shaft. The cells remain stable for a long time after reaction and can also be left overnight and then measured.
- After cooling do not rock the cell until the measurement takes place in order not to suspend the solids that formed during the reaction. Suspended matter disturbs the photometric measurement.




| | |
|-------------------------|-------------------------|
| WTW model no.: | COD1 TC (LR) |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0 - 150 mg/l COD |

Note:


Before using the test with your photometer for the first time, determine the reagent blank value.



Shake the reaction cell so that sediment is suspended.




Carefully pipette 2.0 ml of sample into the cell, close with screw cap and mix vigorously.
Caution, cell becomes very hot!




Heat the cell in the thermoreactor for two hours at 148 °C.




Remove the cell from the thermoreactor and let it cool down in a cell rack.



After approx. 10 min cooling time sway the cell.



Place the cell in the cell rack again and let it cool down **to room temperature**.



Carefully insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- The chloride content of the sample must not exceed 1000 mg/l.
- Homogenize samples containing suspended matter with a disperser.
- Before being inserted in the thermoreactor and for photometric measurements the outside of the cell must be free of any contamination (e.g. fingerprints or drops of water). Wipe the cell with a dry cloth as necessary.
- Let the cell cool down long enough (at least 45 min) before inserting it in the photometer cell shaft. The cells remain stable for a long time after reaction and can also be left overnight and then measured.
- After cooling do not rock the cell until the measurement takes place in order not to suspend the solids that formed during the reaction. Suspended matter disturbs the photometric measurement.




| | |
|-------------------------|-------------------------|
| WTW model no.: | COD2 TC (MR) |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0 - 1500 mg/l COD |

Note:


Before using the test with your photometer for the first time, determine the reagent blank value.



Shake the reaction cell so that sediment is suspended.




Carefully pipette 2.0 ml of sample into the cell, close with screw cap and mix vigorously.
Caution, cell becomes very hot!




Heat the cell in the thermoreactor for two hours at 148 °C.




Remove the cell from the thermoreactor and let it cool down in a cell rack.



After approx. 10 min cooling time sway the cell.



Place the cell in the cell rack again and let it cool down **to room temperature**.



Carefully insert the cell in the photometer cell shaft and start measurement.

Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- The chloride content of the sample must not exceed 1000 mg/l.
- Homogenize samples containing suspended matter with a disperser.
- Before being inserted in the thermoreactor and for photometric measurements the outside of the cell must be free of any contamination (e.g. fingerprints or drops of water). Wipe the cell with a dry cloth as necessary.
- Let the cell cool down long enough (at least 45 min) before inserting it in the photometer cell shaft. The cells remain stable for a long time after reaction and can also be left overnight and then measured.
- After cooling do not rock the cell until the measurement takes place in order not to suspend the solids that formed during the reaction. Suspended matter disturbs the photometric measurement.


Coloration at 435 nm (FB436)

Program no.


43




| | |
|-------------------------|----------------------------|
| WTW model no.: | FB436 |
| Category: | Reagent free test |
| Cell: | 28 mm |
| Measuring range: | 0.5 - 50.0 m ⁻¹ |



Filter the sample as necessary.



Pipette 10.0 ml of sample into the empty cell.



Insert the cell in the photometer cell shaft and start measurement.

Note:

Filtered samples develop the real coloration, unfiltered samples the apparent coloration.


Coloration at 517 nm (FB517)

Program no.


44




| | |
|------------------|----------------------------|
| WTW model no.: | FB517 |
| Category: | Reagent free test |
| Cell: | 28 mm |
| Measuring range: | 0.5 - 50.0 m ⁻¹ |



Filter the sample as necessary.



Pipette 10.0 ml of sample into the empty cell.



Insert the cell in the photometer cell shaft and start measurement.

Note:

Filtered samples develop the real coloration, unfiltered samples the apparent coloration.


Coloration at 610 nm (FB610)

Program no.


45




| | |
|------------------|----------------------------|
| WTW model no.: | FB610 |
| Category: | Reagent free test |
| Cell: | 28 mm |
| Measuring range: | 0.5 - 50.0 m ⁻¹ |



Filter the sample as necessary.



Pipette 10.0 ml of sample into the empty cell.




Insert the cell in the photometer cell shaft and start measurement.

Note:

Filtered samples develop the real coloration, unfiltered samples the apparent coloration.




| | |
|------------------|----------------------------|
| WTW model no.: | 14553 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.05 - 7.50 mg/l Cu |
| | Display in mmol/l possible |



Check the pH value of the sample.
Desired range: pH 4-10.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 5.0 ml of sample into a reaction cell and mix.



Add 5 drops of **Cu-1K**, close cell with screw cap and mix.



Allow to react for 5 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 14767 |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.04 - 3.50 mg/l Cu |
| | Display in mmol/l possible |




Pipette 10.0 ml of sample into the empty cell.




Add 2 green measurers of **Cu-1** and dissolve solids.




Check the pH value. Desired range: pH 7.0-9.5. Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.



Add 10 drops of **Cu-2**, close cell with screw cap and mix.



Allow to react for 5 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value for each test set package started.
- The test sample should be blue. Very high copper concentrations in the sample cause turquoise solutions and too low measured values. Dilute the sample in this case.



| | |
|-------------------------|----------------------------|
| WTW model no.: | 14767 |
| Category: | RT (reagent test) |
| Cell: | 16 mm |
| Measuring range: | 0.10 - 6.00 mg/l Cu |
| | Display in mmol/l possible |




Pipette 5.0 ml of sample into the empty cell.




Add 1 green measurer of **Cu-1** and dissolve solids.




Check the pH value. Desired range: pH 7.0-9.5. Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.



Add 5 drops of **Cu-2**, close cell with screw cap and mix.



Allow to react for 5 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value for each test set package started.
- The test sample should be blue. Very high copper concentrations in the sample cause turquoise solutions and too low measured values. Dilute the sample in this case.




| | |
|-------------------------|----------------------------|
| WTW model no.: | Cu-1 TP |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.00 - 5.00 mg/l Cu |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 4-6.
Correct with diluted sodium hydroxide solution or caustic potash solution as necessary.




Pipette 10.0 ml of sample into the empty cell.




Add the contents of a **VARIO Cu1 F10** powder pack and close the cell with the screw cap.



Mix the contents by carefully swaying the cell (10 x).
Undissolved powder has no adverse effect on measurement.



Allow to react for 2 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value for each test set package started.




| | |
|------------------|----------------------------|
| WTW model no.: | 14561 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.01 - 0.30 mg/l CN |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 1-10.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 5.0 ml of sample into a reaction cell and dissolve solids.




Add 1 level blue microspoon of **CN-3K** and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids. A small amount of solids may remain undissolved.



Allow to react for 10 minutes.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 14557 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.04 - 1.00 mg/l F |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Check the pH value of the sample.
Desired range: pH 3-8.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 5.0 ml of sample into a reaction cell and mix.




Add 1 dose of **F-1K** with the blue measurer and close the cell with the screw cap.




Shake the cell vigorously to dissolve solids.



Allow to react for 5 minutes.



Sway the reaction cell once again.



Insert the cell in the photometer cell shaft and start measurement.

Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 14500 |
| Category: | KT (Reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.10 - 7.00 mg/l HCHO |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Add 1 level green microspoon of **HCHO-1K** and close the cell with the screw cap.




Shake the cell vigorously to dissolve solids.



With a pipette, carefully add 2.0 ml sample, close the cell with the screw cap and mix.
Caution, cell becomes very hot!



Allow to react for 5 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|-------------------------|----------------------------|
| WTW model no.: | 14821 |
| Category: | RT (reagent test) |
| Cell: | 16 mm |
| Measuring range: | 0.5 - 9.0 mg/l Au |
| | Display in mmol/l possible |




Pipette 2.0 ml of the sample into a test tube with screw cap.




Add 2 drops of **Au-1A** and mix.




Add 4 drops of **Au-2A** and mix.




Add 6 drops of **Au-3A** and mix.




Using a pipette, add 6.0 ml **Au-4A** and close the tube.




Shake the tube vigorously for one minute.




Add 6 drops of **Au-5A** and close the tube.




Shake the tube vigorously for one minute.



Aspirate the clear upper phase using a Pasteur pipette.



Fill the clean solution into an empty pipette.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value for each test set package started.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 14549 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.05 - 3.00 mg/l Fe |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 1-10.
Correct with diluted hydrochloric acid as necessary.




Pipette 5.0 ml of sample into a reaction cell and mix.




Add 1 level blue microspoon of **Fe-1K** and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 3 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 14761 |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.05 - 1.50 mg/l Fe |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 1-10.
Correct with diluted hydrochloric acid as necessary.




Pipette 10 ml sample into the empty cell.



Add 6 drops of **Fe-1**, close the cell with the screw cap and mix.



Allow to react for 3 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value for each test set package started.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 14761 |
| Category: | RT (reagent test) |
| Cell: | 16 mm |
| Measuring range: | 0.10 - 3.00 mg/l Fe |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 1-10.
Correct with diluted hydrochloric acid as necessary.




Pipette 5.0 ml of sample into the empty cell.



Add 3 drops of **Fe-1**, close the cell with the screw cap and mix.



Allow to react for 3 minutes.



Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value for each test set package started.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 14896 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 1,0 - 50,0 mg/l Fe |
| | Display in mmol/l possible |

Note: Before using the test with your photometer for the first time, determine the reagent blank value.


Determination of iron(II):




Check the pH value of the sample.
Desired range: pH 3-8.
Correct with diluted sodium hydroxide solution or hydrochloric acid as necessary.



Add 1.0 ml sample with a pipette, close the cell with the screw cap and mix.




Allow to react for 5 minutes.




Insert the cell in the photometer cell shaft and start measurement.


Determination of the sum of iron(II) + iron(III):




Check the pH value of the sample.
Desired range: pH 3-8.
Correct with diluted sodium hydroxide solution or hydrofluoric acid as necessary.




Add 1.0 ml sample with a pipette, close the cell with the screw cap and mix.




Add 1 dose of **Fe-1K** with the blue measurer and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 5 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Note:


For the determination of **total iron** a pretreatment with Crack Set 10C, Cat.No. 252033, or Crack Set 10, Cat.No. 250496 and thermoreactor is necessary.




| | |
|------------------|----------------------------|
| WTW model no.: | Fe-2 TP |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.00 - 3.00 mg/l Fe |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 3-5.
Correct with diluted sodium hydroxide solution or hydrochloric acid as necessary.




Pipette 10.0 ml of sample into the empty cell.




Add the contents of a **VARIO Ferro F10** powder pack and close the cell with the screw cap.



Mix the contents by carefully swaying the cell (10 x).
Undissolved powder has no adverse effect on measurement.



Allow to react for 3 minutes (reaction time).




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- This method covers all forms of dissolved iron and most forms of undissolved iron.
- In the case of samples with visible rust the reaction time should be extended to at least 5 minutes.




| | |
|------------------|----------------------------|
| WTW model no.: | Fe-1 TP |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.00 - 1.80 mg/l Fe |
| | Display in mmol/l possible |




Pipette 10.0 ml of sample into the empty cell.




Add the contents of a **VARIO Iron TPTZ F10** powder pack and close the cell with the screw cap.



To dissolve solids, shake the cell vigorously for approx. 30 seconds.



Allow to react for 3 minutes (reaction time).



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.




| | |
|------------------|----------------------------|
| WTW model no.: | 09717 |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.01 - 4.00 mg/l Pb |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Check the pH value of the sample.
Desired range: pH 3-6.
Correct with diluted nitric acid or ammonia solution as necessary.




Pipette 0.50 ml of **Pb-1** into the empty cell.



Add 0.50 ml **Pb-2** with a pipette and mix.



Add 8.0 ml sample with a pipette, close the cell with the screw cap and mix.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value for each test set package started.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 09717 |
| Category: | RT (reagent test) |
| Cell: | 16 mm |
| Measuring range: | 0.02 - 5.00 mg/l Pb |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.



Check the pH value of the sample.
Desired range: pH 3-6.
Correct with diluted nitric acid or ammonia solution as necessary.




Pipette 0.50 ml of **Pb-1** into the empty cell.



Add 0.50 ml **Pb-2** with a pipette and mix.



Add 8.0 ml sample with a pipette, close the cell with the screw cap and mix.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value for each test set package started.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 00815 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 5.0 - 75.0 mg/l Mg |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Check the pH value of the sample.
Desired range: pH 3-9.
Correct with diluted sodium hydroxide solution or hydrochloric acid as necessary.




Pipette 1.0 ml of sample into a reaction cell and mix.




Add 1.0 ml **Mg-1K** with a pipette, close the cell with the screw cap and mix.



Allow to react for 3 exactly minutes.



Add 3 drops of **Mg-2K**, close cell with screw cap and mix.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|-------------------------|----------------------------|
| WTW model no.: | 00816 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.10 - 5.00 mg/l Mn |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 2-7.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 7.0 ml of sample into a reaction cell and mix.



Add 2 drops of **Mn-1K** and mix.




Allow to react for 2 minutes.



Add 3 drops of **Mn-2K**, close cell with screw cap and mix.



Allow to react for 5 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|-------------------------|----------------------------|
| WTW model no.: | 14770 |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.02 - 5.00 mg/l Mn |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 2-7.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 10.0 ml of sample into the empty cell.




Add 8 drops of **Mn-1** and mix.



Add 4 drops of **Mn-2** and mix.




Allow to react for 2 minutes.



Add 4 drops of **Mn-3**, close the cell with the screw cap and mix.



Allow to react for 2 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value for each test set package started.
- For further notes please refer to the package insert of the test.




| | |
|-------------------------|----------------------------|
| WTW model no.: | 14770 |
| Category: | RT (reagent test) |
| Cell: | 16 mm |
| Measuring range: | 0.04 - 9.00 mg/l Mn |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 2-7.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 10.0 ml of sample into the empty cell.




Add 8 drops of **Mn-1** and mix.



Add 4 drops of **Mn-2** and mix.




Allow to react for 2 minutes.



Add 4 drops of **Mn-3**, close the cell with the screw cap and mix.



Allow to react for 2 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value for each test set package started.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | Mn-1 TP |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.0 - 20.0 mg/l Mn |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 4-5.
Correct with diluted nitric acid or sodium hydroxide solution as necessary.



Pipette 10.0 ml of sample into the empty cell.




Add the contents of a **VARIO MANGANESE Citrate Buffer F10** powder pack and close the cell with the screw cap.



Mix the contents by carefully swaying the cell (10 x).




Add the contents of a **VARIO Sodium Periodate F10** powder pack and close the cell with the screw cap.



Mix the contents by carefully swaying the cell (10 x).



Allow to react for 2 minutes.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value for each test set package started.



| | |
|------------------|----------------------------|
| WTW model no.: | 00860 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.02 - 1.00 mg/l Mo |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Check the pH value of the sample.

Required range:
pH 1-10.


Correct with diluted
sodium hydroxide
solution or sulfuric acid
as necessary.




Add 2 drops of **Mo-1K**
into a reaction cell, close
the cell with the screw
cap and mix.




Add 10.0 ml sample with
a pipette and close the
cell.



Shake the cell vigorously
to dissolve solids.



Allow to react for
2 minutes.




Insert the cell in the
photometer cell shaft
and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|-------------------------|----------------------------|
| WTW model no.: | Mo-1 TP |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.0 - 35.0 mg/l Mn |
| | Display in mmol/l possible |




Pipette 10.0 ml of sample into the empty cell.




Add the contents of a **MolyVer 1 Reagenz** powder pack and close the cell with the screw cap.




Mix the contents by carefully swaying the cell (10 x).




Add the contents of a **MolyVer 2 Reagenz** powder pack and close the cell with the screw cap.



Mix the contents by carefully swaying the cell (10 x).




Add the contents of a **MolyVer 3 Reagenz** powder pack and close the cell with the screw cap.



Mix the contents by carefully swaying the cell (10 x). Undissolved powder has no adverse effect on measurement.



Allow to react for 5 minutes.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.




| | |
|------------------|----------------------------|
| WTW model no.: | 14554 |
| Category: | KT (Reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.10 - 6.00 mg/l Ni |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Check the pH value of the sample.
Required range: pH 3-8.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 5.0 ml of sample into a reaction cell, close the cell with the screw cap and mix.




Allow to react for 1 minute.




Add 2 drops of **Ni-1K**, close the cell with the screw cap and mix.



Add 2 drops of **Ni-2K**, close the cell with the screw cap and mix.



Allow to react for 2 minutes.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 14785 |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.10 - 3.80 mg/l Ni |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Check the pH value of the sample.
Required range: pH 3-8.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 10.0 ml of sample into the empty cell.




Add 2 drops of **Ni-1** and mix. If the solution discolors, go on adding **Ni-1** drop by drop until a slight brown coloring is maintained.



Allow to react for 1 minute.




Add 4 drops of **Ni-2** and mix.




Check the pH value.
Required range: pH 10-12.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.



Add 4 drops of **Ni-3** and mix.



Allow to react for 2 minutes.



Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|------------------------------------|
| WTW model no.: | 14542 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.5 - 14.5 mg/l NO ₃ -N |
| | 2.2 - 64.2 mg/l NO ₃ |
| | Display in mmol/l possible |




Add 1 level blue microspoon of **NO₃-1K** into a reaction cell and close the cell with the screw cap.




To dissolve solids, shake the cell **vigorously for 1 minute**.



Carefully add 1.5 ml of sample with a pipette, close the cell with the screw cap and mix.
Caution, cell becomes very hot!



Allow to react for 10 minutes.



Insert the cell in the photometer cell shaft and start measurement.

Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-------------------------------------|
| WTW model no.: | 14556 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.10 - 2.70 mg/l NO ₃ -N |
| | 0.44 - 11.95 mg/l NO ₃ |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Pipette 2.0 ml of sample into a reaction cell.
Do not mix the contents!




Add 1 level blue microspoon of NO₃-1K. **Immediately** close the cell tightly.
Caution, the contents will foam very much! Use protective goggles and gloves.



To dissolve solids, shake the cell **vigorously for 5 seconds**.



Allow to react for 30 minutes.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|------------------------------------|
| WTW model no.: | 14942 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.2 - 13.0 mg/l NO ₃ -N |
| | 0.9 - 57.5 mg/l NO ₃ |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Pipette 5.0 ml of **NO₃-1** into the empty cell.




Add 1.0 ml sample with a pipette.
Caution, cell becomes very hot!



Immediately add 1.5 ml **NO₃-2** with a pipette and close the cell with the screw cap.




Shake the cell vigorously.




Allow to react for 15 minutes.




Add two level gray microspoons of **NO₃-3** and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 60 minutes.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|------------------------------------|
| WTW model no.: | NO3-1 TC |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.0 - 30.0 mg/l NO ₃ -N |
| | 0.0 - 132.8 mg/l NO ₃ |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Pipette 1.0 ml of sample into a reaction cell and close the cell with the screw cap.




Mix the contents by carefully swaying the cell (10 x).




Add the contents of a **VARIO Nitrate Chromotropic** powder pack and close the cell with the screw cap.



Mix the contents by carefully swaying the cell (10 x). A small amount of solids may remain undissolved.



Allow to react for 5 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.




| | |
|------------------|-------------------------------------|
| WTW model no.: | 14776 |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.01 - 0.30 mg/l NO ₂ -N |
| | 0.03 - 0.99 mg/l NO ₂ |
| | Display in mmol/l possible |




Pipette 10.0 ml of sample into the empty cell.




Add 2 level blue microspoons of **NO₂-1** and close the cell with the screw cap.




Shake the cell vigorously to dissolve solids.



Check the pH value. Desired range: pH 2.0-2.5. Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.



Allow to react for 10 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value for each test set package started.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-------------------------------------|
| WTW model no.: | 14776 |
| Category: | RT (reagent test) |
| Cell: | 16 mm |
| Measuring range: | 0.02 - 0.50 mg/l NO ₂ -N |
| | 0.06 - 1.64 mg/l NO ₂ |
| | Display in mmol/l possible |




Pipette 10.0 ml of sample into the empty cell.




Add 2 level blue microspoons of **NO₂-1** and close the cell with the screw cap.




Shake the cell vigorously to dissolve solids.



Check the pH value. Desired range: pH 2.0-2.5. Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.



Allow to react for 10 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value for each test set package started.
- For further notes please refer to the package insert of the test.




| | |
|-------------------------|-------------------------------------|
| WTW model no.: | N4/25 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.02 - 0.55 mg/l NO ₂ -N |
| | 0.07 - 1.81 mg/l NO ₂ |
| | Display in mmol/l possible |




Pipette 4.0 ml of sample into a reaction cell and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 10 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|-------------------------|---------------------------------------|
| WTW model no.: | N5/25 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.020 - 0.550 mg/l NO ₂ -N |
| | 0.070 - 1.810 mg/l NO ₂ |
| | Display in mmol/l possible |




Pipette 5.0 ml of sample into a reaction cell and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 10 minutes.



Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|-------------------------|---------------------------------------|
| WTW model no.: | 14547 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.020 - 0.550 mg/l NO ₂ -N |
| | 0.070 - 1.810 mg/l NO ₂ |
| | Display in mmol/l possible |




Pipette 5.0 ml of sample into a reaction cell and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 10 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-------------------------------------|
| WTW model no.: | NO2-2 TC |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.30 - 3.00 mg/l NO ₂ -N |
| | 0.99 - 9.85 mg/l NO ₂ |
| | Display in mmol/l possible |




Pipette 0.5 ml of sample into a reaction cell.




Mix the contents by carefully swaying the cell.




Add 1 level black measuring spoon no. 8 of **Nitrit-101** and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 10 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- Store the closed reagents at a temperature of +4 to +8 °C.




| | |
|------------------|-------------------------------------|
| WTW model no.: | NO2-2 TC |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.03 - 0.60 mg/l NO ₂ -N |
| | 0.10 - 1.97 mg/l NO ₂ |
| | Display in mmol/l possible |




Pipette 2.0 ml of sample into a reaction cell.




Mix the contents by carefully swaying the cell.




Add 1 level black measuring spoon no. 8 of **Nitrit-101** and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 10 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- Store the closed reagents at a temperature of +4 to +8 °C.




| | |
|------------------|-------------------------------------|
| WTW model no.: | NO2-1 TP |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.00 - 0.33 mg/l NO ₂ -N |
| | 0.00 - 1.08 mg/l NO ₂ |
| | Display in mmol/l possible |




Pipette 10.0 ml of sample into the empty cell.




Add the contents of a **VARIO Nitri 3 F10** powder pack and close the cell with the screw cap.



Shake the cell. Undissolved powder has no adverse effect on measurement.



Allow to react for 15 minutes.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value for each test set package started.




| | |
|------------------|-------------------------|
| WTW model no.: | 14537 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.50 - 15.00 mg/l N |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Pipette 10.0 ml of sample into an empty cell.



Add 1 level blue microspoon of **N-1K**.




Add 6 drops of **N-2K**, close cell with screw cap and mix.




Heat the cell in the thermoreactor for one hour at 120 °C.



Place the cell in the cell rack again and let it cool down to room temperature (=prepared sample).




Add 1 level blue microspoon of **N-3K** into a reaction cell and close the cell with the screw cap.




To dissolve solids, shake the cell **vigorously for 1 minute**.



With a pipette add 1.5 ml of prepared sample very slowly, close with screw cap and mix vigorously. **Caution, cell becomes very hot!**



Allow to react for 10 minutes.



Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- Sample solution and reagents must have a temperature of 20-25 °C. Temper as necessary.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-------------------------|
| WTW model no.: | Ntot2 TC (HR) |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 10 - 150 mg/l N |

Note:


Before using the test with your photometer for the first time, determine the reagent blank value.



Add the contents of a **VARIO TN Persulfate Rgt.** powder pack into a **TN Hydroxide HR** digestion cell.




With a pipette add 0.5 ml of sample, close the cell with the screw cap and mix vigorously for at least 30 s. A small amount of solids may remain undissolved.




Heat the cell in the thermoreactor at 120 °C for 30 minutes.




Remove the cell from the thermoreactor and let it cool down in a cell rack.



Add the contents of a **VARIO TN Reagent A** powder pack, close the cell with the screw cap and mix for at least 15 s.



Allow to react for 3 minutes.




Add the contents of a **VARIO TN Reagent B** powder pack, close the cell with the screw cap and mix for at least 15 s.



Allow to react for 2 minutes.




Pipette 2.0 ml of prepared sample into a **TN Acid LR/HR (Reagent C)** reaction cell and close the cell with the screw cap.



Mix the contents by carefully swaying the cell (10 x / for approx. 30 s altogether). **Caution, the cell becomes warm!**



Allow to react for 5 minutes.




Insert the cell in the photometer cell shaft and start measurement.




| | |
|------------------|-------------------------|
| WTW model no.: | Ntot1 TC (LR) |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.0 - 25.0 mg/l N |

Note:


Before using the test with your photometer for the first time, determine the reagent blank value.



Add the contents of a **VARIO TN Persulfate Rgt.** powder pack into a **TN Hydroxide LR** digestion cell.




With a pipette add 2.0 ml of sample, close the cell with the screw cap and mix vigorously for at least 30 s. A small amount of solids may remain undissolved.




Heat the cell in the thermoreactor at 120 °C for 30 minutes.




Remove the cell from the thermoreactor and let it cool down in a cell rack.



Add the contents of a **VARIO TN Reagent A** powder pack, close the cell with the screw cap and mix for at least 15 s.



Allow to react for 3 minutes.




Add the contents of a **VARIO TN Reagent B** powder pack, close the cell with the screw cap and mix for at least 15 s.



Allow to react for 2 minutes.




Pipette 2.0 ml of prepared sample into a **TN Acid LR/HR (Reagent C)** reaction cell and close the cell with the screw cap.



Mix the contents by carefully swaying the cell (10 x / for approx. 30 s altogether). **Caution, the cell becomes warm!**




Allow to react for 5 minutes.




Insert the cell in the photometer cell shaft and start measurement.




| | |
|------------------|---------------------------------|
| WTW model no.: | 00607 |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.01 - 1.80 mg/l O ₃ |
| | Display in mmol/l possible |



Check the pH value of the sample.
Desired range: pH 4-8.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 10.0 ml sample into the empty cell.



Add 2 drops of O₃-1 and mix.




Add one level blue microspoon of O₃-2 and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 3 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|-------------------------|---------------------------------|
| WTW model no.: | 00607 |
| Category: | RT (reagent test) |
| Cell: | 16 mm |
| Measuring range: | 0.01 - 3.50 mg/l O ₃ |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 4-8.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 10.0 ml sample into the empty cell.




Add 2 drops of O₃-1 and mix.




Add one level blue microspoon of O₃-2 and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 3 minutes.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|-------------------------|---|
| WTW model no.: | 14551 |
| Category: | KT (Reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.10 - 2.50 mg/l C ₆ H ₅ OH |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Check the pH value of the sample.
Required range:
pH 2-11.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.



Pipette 10 ml of sample into a reaction cell, close the cell with the screw cap and mix.




Add 1 level gray microspoon of **Ph-1K** and close the cell with the screw cap.




Shake the cell vigorously to dissolve solids.




Add 1 level green microspoon of **Ph-2K** and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 1 minute.



Insert the cell in the photometer cell shaft and start measurement.

Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|------------------------------------|
| WTW model no.: | 14546 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.5 - 25.0 mg/l PO ₄ -P |
| | 1.5 - 76.7 mg/l PO ₄ |
| | Display in mmol/l possible |

Note: Before using the test with your photometer for the first time, determine the reagent blank value.



Pipette 5.0 ml of sample into a reaction cell, close the cell with the screw cap and mix.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- This test covers orthophosphate only.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-------------------------------------|
| WTW model no.: | 14848 |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.20 - 1.50 mg/l PO ₄ -P |
| | 0.61 - 4.60 mg/l PO ₄ |
| | Display in mmol/l possible |



Check the pH value of the sample.
Desired range: pH 0-10.
Correct with diluted hydrochloric acid as necessary.




Pipette 10.0 ml of sample into the empty cell.



Add 10 drops of **PO₄-1** and mix.




Add 2 level blue microspoons of **PO₄-2** and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 5 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value for each test set package started.
- This test determines orthophosphate only.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-------------------------------------|
| WTW model no.: | 14848 |
| Category: | RT (reagent test) |
| Cell: | 16 mm |
| Measuring range: | 0.40 - 2.50 mg/l PO ₄ -P |
| | 1.23 - 7.67 mg/l PO ₄ |
| | Display in mmol/l possible |



Check the pH value of the sample.
Desired range: pH 0-10.
Correct with diluted hydrochloric acid as necessary.




Pipette 5.0 ml of sample into the empty cell.



Add 5 drops of **PO₄-1** and mix.




Add 1 level blue microspoon of **PO₄-2** and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 5 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value for each test set package started.
- This test determines orthophosphate only.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-------------------------------------|
| WTW model no.: | P6/25 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.05 - 3.00 mg/l PO ₄ -P |
| | 0.15 - 9.20 mg/l PO ₄ |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 0-10.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.



Pipette 5.0 ml of sample into a reaction cell and mix.




Add one dose of **P-1K** with the green measurer and close the cell with the screw cap.




Heat the reaction cell in the thermoreactor at 120 °C for 30 minutes.




Remove the reaction cell from the thermoreactor and let it cool down to room temperature in a cell rack.




Add 5 drops of **P-2K**, close cell with screw cap and mix.




Add one dose of **P-3K** with the blue measurer and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 5 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- In the case high levels of chloride it is recommended to swap the order of the reagents P-2K and P-3K.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-------------------------------------|
| WTW model no.: | P6/25 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.05 - 3.00 mg/l PO ₄ -P |
| | 0.15 - 9.20 mg/l PO ₄ |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 0-10.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 5.0 ml of sample into a reaction cell and mix.



Add 5 drops of **P-2K**, close cell with screw cap and mix.




Add one dose of **P-3K** with the blue measurer and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 5 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- In the case high levels of chloride it is recommended to swap the order of the reagents P-2K and P-3K.
- For further notes please refer to the package insert of the test.




| | |
|------------------|------------------------------------|
| WTW model no.: | P7/25 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.5 - 15.0 mg/l PO ₄ -P |
| | 1.5 - 46.0 mg/l PO ₄ |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 0-10.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.



Pipette 1.0 ml of sample into a reaction cell and mix.




Add one dose of **P-1K** with the green measurer and close the cell with the screw cap.




Heat the reaction cell in the thermoreactor at 120 °C for 30 minutes.




Remove the reaction cell from the thermoreactor and let it cool down to room temperature in a cell rack.



Add 5 drops of **P-2K**, close cell with screw cap and mix.




Add one dose of **P-3K** with the blue measurer and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 5 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- In the case high levels of chloride it is recommended to swap the order of the reagents P-2K and P-3K.
- For further notes please refer to the package insert of the test.




| | |
|------------------|------------------------------------|
| WTW model no.: | P7/25 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.5 - 15.0 mg/l PO ₄ -P |
| | 1.5 - 46.0 mg/l PO ₄ |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 0-10.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 1.0 ml of sample into a reaction cell and mix.



Add 5 drops of **P-2K**, close cell with screw cap and mix.




Add one dose of **P-3K** with the blue measurer and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 5 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- In the case high levels of chloride it is recommended to swap the order of the reagents P-2K and P-3K.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-------------------------------------|
| WTW model no.: | PO4-1 TP |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.00 - 0.80 mg/l PO ₄ -P |
| | 0.00 - 2.45 mg/l PO ₄ |
| | Display in mmol/l possible |



Pipette 10.0 ml of sample into the empty cell.




Add the contents of a **VARIO Phos3 F10** powder pack and close the cell with the screw cap.



Shake the cell for 10 to 15 seconds. Undissolved powder has no adverse effect on measurement.



Allow to react for 2 minutes.



Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value for each test set package started.




| | |
|------------------|-------------------------------------|
| WTW model no.: | PO4-2 TC |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.00 - 1.60 mg/l PO ₄ -P |
| | 0.00 - 4.91 mg/l PO ₄ |
| | Display in mmol/l possible |




Pipette 5.0 ml of sample into a reaction cell and close the cell with the screw cap.




Mix the contents by carefully swaying the cell.




Add the contents of a **VARIO Phos 3 F10** powder pack and close the cell with the screw cap.



To dissolve solids, shake the cell for approx. 10 to 15 seconds. A small amount of solids may remain undissolved.



Allow to react for 2 minutes.



Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.




| | |
|------------------|-------------------------------------|
| WTW model no.: | PO4-3 TC |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.00 - 1.10 mg/l PO ₄ -P |
| | 0.00 - 3.37 mg/l PO ₄ |
| | Display in mmol/l possible |




Pipette 5.0 ml of sample into a reaction cell.



Add the contents of a **VARIO Potassium Persulfate F10** powder pack and close the cell with the screw cap.




Shake the cell vigorously to dissolve solids.




Heat the cell in the thermoreactor for 30 minutes at 120 °C.




Remove the cell from the thermoreactor and let it cool down in a cell rack.



With a pipette add 2.0 ml of 1.54 N sodium hydroxide solution, close the cell with the screw cap and mix the contents by carefully swaying the cell.




Add the contents of a **VARIO Phos 3 F10** powder pack and close the cell with the screw cap.



To dissolve solids, shake the cell for approx. 10 to 15 seconds. A small amount of solids may remain undissolved.



Allow to react for 2 minutes.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.




| | |
|------------------|------------------------------------|
| WTW model no.: | 00616 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 1.0 - 70.0 mg/l PO ₄ -P |
| | 3.1 - 214.6 mg/l PO ₄ |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Check the pH value of the sample.
Desired range: pH 0-10.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 0.20 ml of sample into a reaction cell and mix.




Add 5 drops of PO₄-1K, close the cell with the screw cap and mix.




Add 1 dose of PO₄-2K with the blue measurer and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 5 minutes.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- This test covers orthophosphate only.
- For further notes please refer to the package insert of the test.




| | |
|------------------|------------------------------------|
| WTW model no.: | 00798 |
| Category: | RT (reagent test) |
| Cell: | 16 mm |
| Measuring range: | 1.0 - 50.0 mg/l PO ₄ -P |
| | 3.1 - 153.3 mg/l PO ₄ |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.



Check the pH value of the sample.
Desired range: pH 0-10.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 8.0 ml of deionized water into the empty cell.




Add 0.50 ml of sample with a pipette and mix.



Add 0.50 ml **PO₄-1** with a pipette and mix.




Add 1 dose of **PO₄-2K** with the blue measurer and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 5 minutes.



Insert the cell in the photometer cell shaft and start measurement.

Notes:

- We recommend to determine a new reagent blank value for each test set package started.
- This test covers orthophosphate only.
- For further notes please refer to the package insert of the test.


Phosphate: Orthophosphate

Program no.

51




| | |
|------------------|-------------------------------------|
| WTW model no.: | 14543 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.05 - 3.00 mg/l PO ₄ -P |
| | 0.15 - 9.20 mg/l PO ₄ |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 0-10.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 5.0 ml of sample into a reaction cell and mix.



Add 5 drops of **P-2K**, close the cell with the screw cap and mix.




Add 1 dose of **P-3K** with the blue measurer and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 5 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|------------------------------------|
| WTW model no.: | 14729 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.5 - 15.0 mg/l PO ₄ -P |
| | 1.5 - 46.0 mg/l PO ₄ |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 0-10.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 1.0 ml of sample into a reaction cell and mix.



Add 5 drops of **P-2K**, close the cell with the screw cap and mix.




Add 1 dose of **P-3K** with the blue measurer and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 5 minutes.



Insert the cell in the photometer cell shaft and start measurement.

Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.


Phosphate: Total phosphate

Program no.


52




| | |
|------------------|-------------------------------------|
| WTW model no.: | 14543 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.05 - 3.00 mg/l PO ₄ -P |
| | 0.15 - 9.20 mg/l PO ₄ |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 0-10.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.



Pipette 5.0 ml of sample into a reaction cell and mix.




Add 1 dose of **P-1K** with the green measurer and close the cell with the screw cap.




Heat the reaction cell in the thermoreactor at 120 °C for 30 minutes.




Remove the reaction cell from the thermoreactor and let it cool down to room temperature in a cell rack.



Add 5 drops of **P-2K**, close the cell with the screw cap and mix.




Add 1 dose of **P-3K** with the blue measurer and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 5 minutes.



Insert the cell in the photometer cell shaft and start measurement.

Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.


Phosphate: Total phosphate

Program no.


54




| | |
|------------------|------------------------------------|
| WTW model no.: | 14729 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.5 - 15.0 mg/l PO ₄ -P |
| | 1.5 - 46.0 mg/l PO ₄ |
| | Display in mmol/l possible |




Check the pH value of the sample.
Desired range: pH 0-10.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.



Pipette 1.0 ml of sample into a reaction cell and mix.




Add 1 dose of **P-1K** with the green measurer and close the cell with the screw cap.




Heat the reaction cell in the thermoreactor at 120 °C for 30 minutes.




Remove the reaction cell from the thermoreactor and let it cool down to room temperature in a cell rack.




Add 5 drops of **P-2K**, close the cell with the screw cap and mix.




Add 1 dose of **P-3K** with the blue measurer and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 5 minutes.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 00615 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 30 - 300 mg/l K |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.



Filter turbid sample solutions.




Check the pH value of the sample.
Desired range: pH 3-12.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 0.50 ml of sample into a reaction cell, close the cell with the screw cap and mix.




Check the pH value of the sample.
Desired range: pH 10,0-11,5.




Add 6 drops of **K-1K**, close cell with screw cap and mix.




Add 1 level blue microspoon of **K-2K** and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for exactly 5 minutes. **Then measure immediately.**



Insert the cell in the photometer cell shaft and start measurement.

Notes:


- The turbidity of the measurement solution remains stable for only a short time (the measurement value increases by 5 to 7 % per minute).
- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 14562 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 5.00 - 50,00 mg/l K |
| | Display in mmol/l possible |

Note:

Before using the test with your photometer for the first time, determine the reagent blank value.



Filter turbid sample solutions.




Check the pH value of the sample.
Desired range: pH 3-12.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Pipette 2.0 ml of sample into a reaction cell, close the cell with the screw cap and mix.




Check the pH value of the sample.
Desired range: pH 10.0-11.5.




Add 6 drops of **K-1K**, close the cell with the screw cap and mix.




Add 1 level blue microspoon of **K-2K** and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for exactly 5 minutes. **Then measure immediately.**




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- The turbidity of the measurement solution remains stable for only a short time (the measurement value increases by 5 to 7 % per minute).
- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|-------------------------|--------------------------------|
| WTW model no.: | Si-2 TP (HR) |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.0 - 70.0 mg/l SiO_2 |
| | 0.0 - 32.7 mg/l Si |
| | Display in mmol/l possible |




Pipette 10.0 ml of sample into the empty cell.




Add the contents of a **VARIO Silica HR Molybdate F10** powder pack and close the cell with the screw cap.




Mix the contents by carefully swaying the cell.




Add the contents of a **VARIO Silica HR Acid Rgt F10** powder pack and close the cell with the screw cap.




Mix the contents by carefully swaying the cell.




Allow to react for 10 minutes.




Add the contents of a **VARIO Silica Citric Acid F10** powder pack and close the cell with the screw cap.



Mix the contents by carefully swaying the cell.



Allow to react for 2 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value for each test set package started.
- The sample temperature has to be between 15 and 25 °C.




| | |
|-------------------------|-------------------------------|
| WTW model no.: | Si-2 TP (HR) |
| Category: | RT (reagent test) |
| Cell: | 16 mm |
| Measuring range: | 0 - 100 mg/l SiO ₂ |
| | 0 - 46 mg/l Si |
| | Display in mmol/l possible |




Pipette 10.0 ml of sample into the empty cell.




Add the contents of a **VARIO Silica HR Molybdate F10** powder pack and close the cell with the screw cap.




Mix the contents by carefully swaying the cell.




Add the contents of a **VARIO Silica HR Acid Rgt F10** powder pack and close the cell with the screw cap.




Mix the contents by carefully swaying the cell.




Allow to react for 10 minutes.




Add the contents of a **VARIO Silica Citric Acid F10** powder pack and close the cell with the screw cap.



Mix the contents by carefully swaying the cell.



Allow to react for 2 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value for each test set package started.
- The sample temperature has to be between 15 and 25 °C.




| | |
|-------------------------|-----------------------------------|
| WTW model no.: | Si-1 TP (LR) |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.00 - 1.60 mg/l SiO ₂ |
| | 0.00 - 0.75 mg/l Si |
| | Display in mmol/l possible |



Pipette 10.0 ml of sample into the empty cell.



Add 15 drops of **Vario Molybdate 3 Reagent Solution** and close the cell with the screw cap.




Mix the contents by carefully swaying the cell.




Allow to react for 4 minutes (temperature dependence, see note).




Add the contents of a **VARIO Silica Citric Acid F10** powder pack and close the cell with the screw cap.



Mix the contents by carefully swaying the cell.




Allow to react for 1 minute (temperature dependence, see note).



Add the contents of a **VARIO LR Silica Amino Acid F F10** powder pack and close the cell with the screw cap.



Allow to react for 2 minutes. If SiO₂ is present in the sample, the solution becomes blue.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value for each test set package started.
- The above mentioned reaction times are valid at room temperature (20 °C). At 10 °C the reaction time has to be doubled, at 30 °C it has to be halved.




| | |
|-------------------------|------------------------------------|
| WTW model no.: | 14794 |
| Category: | RT (reagent test) |
| Cell: | 16 mm |
| Measuring range: | 0.10 - 5.00 mg/l Si |
| | 0.21 - 10.70 mg/l SiO ₂ |
| | Display in mmol/l possible |




Pipette 5.0 ml of sample into the empty cell.




Add 3 drops of **Si-1** and mix.




Check the pH value of the sample.
Desired range:
pH 1.2-1.6.
If necessary, adjust by adding more drops of **Si-1**.




Allow to react for 3 minutes.




Add 3 drops of **Si-2** and mix.



Add 0.50 ml **Si-3** with a pipette, close the cell with the screw cap and mix.



Allow to react for 10 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- Clean the cells and all glass vessels that sometimes come into contact with the blue complex as follows: Fill the vessels with sodium hydroxide solution (approx. 0.4 %) and let it act for max. 1 hour.
- For further notes please refer to the package insert of the test.




| | |
|-------------------------|-----------------------------------|
| WTW model no.: | 14794 |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0.05 - 2.50 mg/l Si |
| | 0.11 - 5.35 mg/l SiO ₂ |
| | Display in mmol/l possible |




Pipette 10.0 ml of sample into the empty cell.




Add 6 drops of **Si-1** and mix.




Check the pH value of the sample.
Desired range:
pH 1.2-1.6.
If necessary, adjust by adding more drops of **Si-1**.




Allow to react for 3 minutes.




Add 6 drops of **Si-2** and mix.



Add 1.00 ml **Si-3** with a pipette, close the cell with the screw cap and mix.



Allow to react for 10 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- Clean the cells and all glass vessels that sometimes come into contact with the blue complex as follows: Fill the vessels with sodium hydroxide solution (approx. 0.4 %) and let it act for max. 1 hour.
- For further notes please refer to the package insert of the test.




| | |
|-------------------------|-----------------------------------|
| WTW model no.: | 00857 |
| Category: | RT (reagent test) |
| Cell: | 16 mm |
| Measuring range: | 0.5 - 50.0 mg/l Si |
| | 1.1 - 106.9 mg/l SiO ₂ |
| | Display in mmol/l possible |



Pipette 4.0 ml of sample into the empty cell.




Add 4 drops of **Si-1** and mix.



Add 2.0 ml **Si-2** with a pipette and mix.




Allow to react for 2 minutes.



Add 4 drops of **Si-3** and mix.



Allow to react for 2 minutes.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- Clean the cells and all glass vessels that sometimes come into contact with the blue complex as follows: Fill the vessels with sodium hydroxide solution (approx. 0.4 %) and let it act for max. 1 hour.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 14831 |
| Category: | RT (reagent test) |
| Cell: | 16 mm |
| Measuring range: | 0.25 - 2.75 mg/l Ag |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Pipette 10.0 ml of sample into the empty cell.



Add 2 drops of **Ag-1**.




Add one level green microspoon of **Ag-2** and close the cell with the screw cap.




Heat the cell in the thermoreactor at 120 °C (100 °C) for 60 minutes.




Remove the cell from the thermoreactor and let it cool down in a cell rack.




Sway the cell before opening it.




Add 3 drops of **Ag-3**, close the cell with the screw cap and mix.



Check the pH value of the solution. Required range: pH 4-10. Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.



Add 1 drop of **Ag-4**, close the cell with the screw cap and mix.




Add 5 drops of **Ag-5**, close the cell with the screw cap and mix.



Add 1.0 ml **Ag-6** with a pipette, close the cell with the screw cap and mix.



Allow to react for 5 minutes.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value for each test set package started.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 00885 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 10 - 300 mg/l Na |
| | Display in mmol/l possible |




Pipette 0.50 ml **Na-1K** into a reaction cell and mix.



Add 0.50 ml of sample, close cell with screw cap and mix.



Allow to react for 1 minute.




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-------------------------------|
| WTW model no.: | 14548 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 25 - 250 mg/l SO ₄ |
| | Display in mmol/l possible |




Filter turbid sample solutions.




Check the pH value of the sample.
Desired range: pH 0-10.
Correct with diluted sodium hydroxide solution or hydrochloric acid as necessary.



Pipette 5.0 ml of sample into a reaction cell and mix.




Add 1 level green microspoon of **SO₄-1K** and close the cell with the screw cap.



Shake the cell vigorously to dissolve solids.



Allow to react for 2 minutes. **Then measure immediately.**




Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value for each test set package started.
- For further notes please refer to the package insert of the test.




| | |
|------------------|-----------------------------|
| WTW model no.: | SO4-1 TP |
| Category: | RT (reagent test) |
| Cell: | 28 mm |
| Measuring range: | 0 - 70 mg/l SO ₄ |
| | Display in mmol/l possible |




Pipette 10.0 ml of sample into the empty cell.




Add the contents of a **VARIO SULPHA 4 F10** powder pack and close the cell with the screw cap.



Mix the contents by swaying. If sulfate is present in the sample, a white turbidity occurs.



Allow to react for 5 minutes.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value for each test set package started.




| | |
|------------------|----------------------------|
| WTW model no.: | 14697 |
| Category: | KT (Reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.05 - 2.00 mg/l MBAS |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Check the pH value of the sample.
Required range:
pH 5-10.
Correct with diluted sodium hydroxide solution or hydrochloric acid as necessary.




Pipette 5.0 ml of sample into a reaction cell.
Do not mix the contents!




Add 3 drops of **T-1K**.
Do not mix the contents!




Add 2 drops of **T-2K**, close the cell with the screw cap and mix.




Shake the cell vigorously for 30 seconds.



Allow to react for 10 minutes.



Sway the cell before measuring.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|------------------------------|
| WTW model no.: | 01787 |
| Category: | KT (Reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.10 - 7.50 mg/l TritonX-100 |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Check the pH value of the sample.
Required range: pH 3-9.
Correct with diluted sodium hydroxide solution or hydrochloric acid as necessary.




Pipette 4.0 ml of sample into a reaction cell and close the cell with the screw cap.




Shake the cell vigorously for 60 seconds.



Allow to react for 2 minutes.



Sway the cell before measuring.



Insert the cell in the photometer cell shaft and start measurement.

Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.

Water hardness, total hardness


Program no.

46




| | |
|------------------|----------------------------|
| WTW model no.: | 00961 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 5 - 215 mg/l GH/Ca |
| | 0.7 - 30.1 °d |
| | 0.9 - 37.6 °e |
| | 1.2 - 53.7 °f |
| | 7 - 301 mg/l CaO |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.




Check the pH value of the sample.
Desired range: pH 3-9.
Correct with diluted sodium hydroxide solution or hydrochloric acid as necessary.




Pipette 1.0 ml of sample into a reaction cell and mix.



Add 1.0 ml **H-1K** with a pipette, close the cell with the screw cap and mix.



Allow to react for 3 minutes.



Insert the cell in the photometer cell shaft and start measurement.

Notes:


- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 14566 |
| Category: | KT (reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.20 - 5.00 mg/l Zn |
| | Display in mmol/l possible |

Note:


Before using the test with your photometer for the first time, determine the reagent blank value.




Check the pH value of the sample.
Desired range: pH 3-10.
Correct with diluted sodium hydroxide solution or sulfuric acid as necessary.




Add 5 drops of **Zn-1K** into a reaction cell, close the cell with the screw cap and mix.




Add 0.50 ml of sample with a pipette, close the cell with the screw cap and mix.



Add 5 drops of **Zn-2K**, close cell with screw cap and mix.



Allow to react for 15 minutes.



Insert the cell in the photometer cell shaft and start measurement.


Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.
- For further notes please refer to the package insert of the test.




| | |
|------------------|----------------------------|
| WTW model no.: | 00861 |
| Category: | KT (Reaction cell test) |
| Cell: | 16 mm |
| Measuring range: | 0.025 - 1.000 mg/l Zn |
| | Display in mmol/l possible |


Note: Before using the test with your photometer for the first time, determine the reagent blank value.



Pipette 10.0 ml of sample into an empty beaker.




Add one level gray microspoon of **Zn-1K** and dissolve the solids (=prepared sample).



Using a pipette, add 0.50 ml of **Zn-2K** into a reaction cell and close the cell with the screw cap.




Carefully add 2.0 ml of the **prepared sample** with a pipette, close the cell with the screw cap and mix.



Add 5 drops of **Zn-3K**, close cell with screw cap and mix.



Allow to react for 15 minutes.



Insert the cell in the photometer cell shaft and start measurement.

Notes:

- We recommend to determine a new reagent blank value each time when starting a new package using a reaction cell from the same package.



Wissenschaftlich-Technische Werkstätten GmbH

Dr.-Karl-Slevogt-Straße 1
D-82362 Weilheim

Germany

Tel: +49 (0) 881 183-0
+49 (0) 881 183-100
Fax: +49 (0) 881 183-420
E-Mail: Info@WTW.com
Internet: <http://www.WTW.com>